

Table S4. The results of intra-group repeatability.

Number of copies/(copies/ μ L)	C _t values			Mean C _t	SD	CV (%)
	1	2	3			
10 ⁸	14.2	14.3	14.3	14.3	0.08	0.5
10 ⁷	17.6	17.5	17.4	17.5	0.06	0.5
10 ⁶	21.6	22.3	21.7	21.9	0.4	1.8
10 ⁵	24.8	25.1	25.1	25.0	0.2	0.8
10 ⁴	27.6	27.8	28.3	27.9	0.3	1.3

C_t: Cycle Threshold; SD: Standard deviation; CV: Coefficient of variation.