

Review

Not peer-reviewed version

Antioxidant Determining Using an Electrochemical Method

[Rani Melati Sukma](#) , [Dyah Iswantini](#) * , Novik Nurhidayat , [Mohamad Rafi](#) , [Dita Ariyanti](#)

Posted Date: 28 July 2023

doi: [10.20944/preprints2023071654.v2](https://doi.org/10.20944/preprints2023071654.v2)

Keywords: antioxidants; biocatalyst; biosensor; electrochemical method; sensor



Preprints.org is a free multidiscipline platform providing preprint service that is dedicated to making early versions of research outputs permanently available and citable. Preprints posted at Preprints.org appear in Web of Science, Crossref, Google Scholar, Scilit, Europe PMC.

Copyright: This is an open access article distributed under the Creative Commons Attribution License which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Review

Antioxidant Determining Using an Electrochemical Method

Rani Melati Sukma ¹, Dyah Iswantini ^{1,3,*}, Novik Nurhidayat ², Mohamad Rafi ^{1,3} and Dita Ariyanti ⁴

¹ Department of Chemistry, Faculty of Mathematics and Natural Sciences, IPB University, Bogor 16680, Indonesia

² Research centre for Applied Microbiology, National Research and Innovation Agency, BRIN, Cibinong 16911, Indonesia

³ Tropical Biopharmaca Research Center, IPB University, Bogor 16128, Indonesia

⁴ Department of Chemistry, Faculty of Military Mathematics and Natural Sciences, Indonesian Defense University, Bogor 16810, Indonesia.

* Correspondence: dyahis@apps.ipb.ac.id; Tel.: +62818745151

Abstract: Antioxidants are very beneficial for health because they protect the body from the effects of free radicals on various degenerative diseases caused by food contamination, air pollution, sunlight, etc. In general, methods for measuring the capacity of antioxidants generally use accurate methods such as spectrophotometry and chromatography. Still, it takes time, sample preparation and must be done in a laboratory with particular expertise. Therefore, a new, more practical method needs to be developed for determining antioxidants, namely the electrochemical method. The electrochemical method is promising to develop because it has several advantages, including high sensitivity and fast response. The electrochemical method discussed in this article reviews sensors, biosensors, and nanosensors. This paper comprehensively analyzes contemporary developments in electrochemical biosensor techniques and antioxidant evaluation methodologies. The discussion centers on utilizing multiple biosensors. Electrochemical biosensors have been determined to be prevalent in analyzing food quality, assessing active factor functionality, and screening practical components. The present study outlines the difficulties linked with electrochemical bio-sensor technology and provides insights into the potential avenues for future research in this domain.

Keywords: antioxidants; biocatalyst; biosensor; electrochemical method; sensor

1. Introduction

Increasing body immunity and maintaining health through antioxidant preparations has become a new concern after the Covid-19 pandemic. Antioxidants are chemical substances that interrupt the cascade of free radical reactions within the human body. Free radicals are generated as byproducts of metabolic and physiological processes and are an essential part of the immune system of aerobic organisms, including humans. According to sources [1,2], radicals have one or more unpaired electrons, rendering them unstable and able to damage other atoms by losing electrons to become stable. Reactive oxygen/nitrogen species (ROS/RNS) are called these substances, such as hydroxyl, hydrogen peroxide, superoxide, nitric oxide, and peroxynitrite [3,4]. Types of antioxidants are divided into two: endogenous antioxidants produced by our bodies [5], and exogenous antioxidants that come from food or nutritional supplements; one example is polyphenolic compounds [6,7]. The high content of phenolic components in the samples resulted in high levels of antioxidant activity [8].

In some diets, spices, herbs, fruits, and vegetables can provide additional natural antioxidants to support antioxidant defenses [9]. Antioxidants sourced from food, beverages, and herbal medicines must be controlled for their quality so that the antioxidants consumed can optimally counteract free radicals in the body to help prevent disease. Several research studies have indicated

that medical conditions, including inflammation, osteoporosis, hepatopathy, diabetes, cancer, and neuro-degenerative diseases, are frequently linked with elevated levels of oxidative stress [10–14]. There are significant differences between RNS and ROS, and how antioxidants protect the body from them. Enhancing cellular defenses through antioxidants has effectively neutralized oxidative stress [15,16].

Spices and herbs are known extensively to have high antioxidant activity and beneficial effects on human health in certain spices. Antioxidants derived from spices include bioactive compounds consisting of flavonoids, phenolic compounds, compounds containing sulfur, tannins, alkaloids, diterpenes, and vitamins [15]. The compounds exhibit variations in their antioxidant efficacy. For instance, flavonoids can eliminate free radicals and establish associations with catalytic metal ions, rendering them reactive. Numerous academic literature has shown that spices and herbs, including but not limited to rosemary, sage, and oregano, possess high levels of phenolic compounds and antioxidants. Antioxidants can safeguard oils against oxidative degradation. Antioxidants, when incorporated into food, can impede the formation of harmful oxidation byproducts, preserve the nutritional tributes, and prolong the duration of product storage. Spices contain inherent antioxidants that aid in the mitigation of oxidative stress. *Oxidative stress* is a biological state that results from heightened concentrations of unpaired electrons, known as free radicals, within cellular and tissue environments. Various detrimental factors, including gamma radiation, UV and X-rays, psychological stress, contaminated food, unfavorable environmental circumstances, strenuous physical exertion, tobacco use, and alcohol addiction, can trigger this condition [17,18].

According to previous studies, the process of inhibiting free radicals can be achieved by antioxidants by donating an electron to oxidant compounds, thereby impeding their activity. The efficacy of antioxidants in mitigating the effects of free radicals can be classified into two distinct groups. The initial classification, namely primary, pertains to antioxidants that undergo chain termination to impede the generation of free radicals. In this process, antioxidants donate hydrogen from their active hydroxyl groups to produce more radicals [19]. The second category pertains to the deactivation of free radicals by transferring single electrons to form more stable substances, accomplished by antioxidants [20]. Antioxidants are compounds that have many benefits, including being used for beauty and cosmetic products [21], health care [22], food ingredients and preservatives [23], production of silver nanoparticles [24], and others.

Antioxidant detection can use spectrophotometry [25], colorimetry [26], chromatography [27], spectroscopy, and electrochemical methods [28]. The utilization of the chromatographic technique entails a higher cost due to the high expense of the equipment involved. While the method can differentiate between distinct antioxidant constituents in various food items, it only furnishes data regarding their concentration. The spectroscopic technique relies on the spectral characteristics of a reference material because of its determinant principle, which causes the measurement results to have unavoidable errors, such as determining the actual color of the sample, such as orange juice, etc. For this reason, it is necessary to pay attention to and develop simple, sensitive, and fast methods for analysis, such as the electrochemical method. In general, the conventional method is sensitive and efficient. Still, the work is usually carried out in a centralized laboratory, requires resources and experts in the field, and requires expensive costs and a long time. The electrochemical approach offers numerous benefits, such as rapid detection time, minimal sample volume requirement, exceptional precision, and heightened sensitivity. By circumventing the need for laborious sample pre-treatment, interference from colored samples can be minimized [29–31]. One method utilized to evaluate antioxidant capacity is the electrochemical approach, which is preferred for its accuracy, affordability, simplicity, rapid response, and high sensitivity [32]. A multitude of electrochemical methodologies, including square wave voltammetry (SWV), cyclic voltammetry (CV), and differential pulse voltammetry (DPV), have been extensively utilized in various research endeavors to explore redox systems and produce results [33,34].

The information literature review was collected from scientific journals, Wiley Online Library, Scopus, Google Scholar, and Science Direct. The keyword is biosensor antioxidant and electrochemical method. The articles obtained are filtered by title, abstract, and full text.

2. Antioxidants

Based on how they react with free radicals, antioxidants are classified as either enzymatic or non-enzymatic. Enzymatic antioxidants remove free radicals by donating electrons to free radical species produced by the body. Meanwhile, non-enzymatic antioxidants remove free radicals by interfering with the chain reaction of free radicals, and many non-enzymatic antioxidants come from food [35]. Enzymatic antioxidants can be classified into two discrete classifications: primary and secondary. The enzyme antioxidant system in the body consists of glutathione peroxidase (GPX), catalase compounds (CAT), and superoxide dismutase (SOD) [36]. Non-enzymatic antioxidants commonly found in crops are polyphenols, for example, flavonoids. Non-enzyme antioxidants function in balancing the health system of organisms. Vitamin E (tocopherol) and ascorbic acid (vitamin C) are non-enzyme antioxidants (Figure 1) [37].




Figure 1. Classification of Antioxidants.

Antioxidants with enzyme-based work by breaking down and removal of free radicals. Generally, antioxidant enzymes remove harmful oxygen from the product and then convert it into H_2O_2 , then H_2O_2 in the process requires several metal cofactors (copper, zinc, manganese, and iron) [38]. Polyphenols are classified into two groups based on their work against free radicals. First, the polyphenol group fights free radicals by breaking down the main chain to reduce or delay free radical production, which will trap the free radicals produced. Free radical scavengers in secondary antioxidants are carried out on the substrate. In contradistinction, primary antioxidants neutralize free radicals through three mechanisms, namely hydrogen atom transfer (HAT), sequential electron transfer proton transfer (SETPT), and sequential proton loss electron transfer (SPLET), by donating hydrogen atoms to peroxy radicals [39].

Flavonoids can stop free radicals by binding free electrons to conjugated hydroxyl species [40]. The chelating properties of flavonoids also encourage inhibiting free radicals in the body [41]. Flavonoids can act as anti-allergic, anti-diabetic, anti-inflammatory, and antioxidant [42]. Phenolic acid is an antioxidant with conjugated aromatic rings and substituted hydroxyl groups. The functional groups' composition and quantity significantly affect the phenolic acid's activity, as

reported in reference [43]. There are several mechanisms of free radical scavenging by phenolic acid, carried out by donating H atoms, donating electrons, and binding of oxidative free radicals [44]. Apart from having many functions for the body, antioxidants also extend the shelf life of food because antioxidants can slow down reactive species, such as ROS and RNS, which can increase the food decay rate [16]. Another function of antioxidants is changing metal peroxide to become more stable and inhibiting lipoxygenase in prooxidative enzymes [45]. The position of the functional group of the active compound is also important, such as OH or NH₂, with the *ortho* position being the most active compared to the *para* and *meta* positions.

3. Electrochemical Sensors

The electrochemical sensor transforms the signal resulting from the chemical interaction between the analyte and the identifying component to obtain quantitative or qualitative information. The sensor signal is then transformed into an analytical signal. The resulting signal can change the voltage, conductance, and current. Electrochemical sensors have many advantages over mass, thermal and optical sensors, and electrochemical sensors demonstrate to possess superior sensitivity, ease of use, cost-effectiveness, and versatility in analyzing analytes in various states, including liquid, solid, and gas. [46]. Electrochemical sensors for determining antioxidants are developed using various electrodes, receptors, and transducers. Several times, nanomaterials use to enhance sensor performance, sensitivity, stability, and selectivity. Antioxidants tend to have significant redox properties, so they can be detected directly [47].

The integration of nanomaterials into electrochemical sensors has yielded significant improvements. Nanomaterials exhibit distinct thermal, mechanical, optical, electrical, and magnetic characteristics contingent upon their size and can be readily tailored by manipulating their shape and dimensions. [48]. The characteristics of the added nanomaterials have a significant impact on improving the sensor's electrochemical performance. [49]. Measurement of the redox potential of antioxidants using the electrochemical sensor method in reporting the results must be included with the reference electrode utilized because it cannot be quantified in absolute terms. Determining antioxidants through electrochemical sensors is significantly impacted by the solvent employed, as the solvent substantially impacts the reactivity of antioxidants. Therefore, it is necessary to consider the redox reactions occurring in several different solvent types to measure solvents' impact on the antioxidant measurement [50]. Measurement of antioxidants with other methods pays little attention to the influence of solvents, electrolytes, radical intermediates, reactant concentrations, and pH. At the same time, the electrochemical sensor method can estimate this well [51]. Several literature studies on sensors in detecting antioxidants using electrochemical methods can be seen in Table 1.

This research concentrated on developing electrochemical sensors for detecting uric acid and gallic acid in green tea and fruit juice samples using carbon paste electrodes modified by ZrO₂ nanocomposites, chlorine chloride, and gold nanoparticles using the DPV electrochemical method. It showed promising results, obtaining a LOD value of 2.5×10^{-8} M and a linear range of 0.22-55 μ M for gallic acid samples, while for uric acid samples, the LOD is 1.5×10^{-8} M and a concentration range of 0.12-55 μ M [52]. Numerous electrochemical sensor techniques have been devised for the quantification of antioxidant activity in food specimens, such as determining the type of antioxidant in the form of phloretin (Ph) in spice samples using the cyclic voltammetry method and glassy carbon electrodes to produce a LOD of 4.1×10^{-6} with a concentration range of 9.9×10^{-6} - 1.07×10^{-4} M [53]. A glassy carbon electrode modified with a hydroxide film based on Zn Al-NO₃ on antioxidant measurements of gallic acid and caffeic acid using the pulse differential voltammetry detection method produces a LOD of 1.6-2.6 μ M [44].

The development of electrochemical sensors using differential pulse and cyclic voltammetry methods with glass carbon working electrodes with in situ activity for the determination of tertiary antioxidant butyl hydroquinone resulted in a detection limit of 67 nM and a linear range of 1.0-1.1 μ m [54]. Electrochemical detection using the differential pulse polarography method with mercury-dropping electrodes in determining gallic acid as an antioxidant resulted in a detection limit of 0.3 μ m with a linear range of 1.0-50 μ m. The total polyphenols measured in red wine samples averaged

1,987 ppm and 238.1 ppm in white wine samples [55]. Development of electrochemical sensors with different measurement methods, namely using an electrochemical detection method in the form of square wave voltammetry with several working electrodes and different samples, firstly using a glassy carbon working electrode immobilized by purine biosensors in measuring beverage samples to determine total antioxidant capacity, secondly, using the SPCE working electrode on an antioxidant in the form of ascorbic acid produced a detection limit of 0.09 nM, melatonin produced a detection limit of 0.04 nM, and N-acetylcysteine produced a detection limit of 0.07 nM [56].

The measurement of chlorogenic acid in nutraceuticals involves using a voltammetry sensor based on an SPCE modified with graphene and gold nanoparticles. This sensor utilizes a glassy carbon electrode modified with 4-methylpyridinium iodine to detect the presence of caffeine acid, an antioxidant. The electrochemical sensor comprises three distinct screen-printed electrodes (SPEs): a carbon-based electrode, a graphene-based electrode, and a graphene-modified electrode that incorporates gold nanoparticles dissolved in various solutions. The lowest LOD and LOQ were obtained by SPE based on graphene (GPH) and gold nanoparticles (GNP) with LOD = 0.62×10^{-7} M and LOQ = 1.97×10^{-7} M. The present study suggests that utilizing GPH-GNP-SPE enhances sensor response in terms of sensitivity and reversibility for the accurate determination of real-world samples. This conclusion is supported by the validation of FTIR results, which indicates no statistically significant differences [57].

Table 1. Some literature on antioxidant detection uses electrochemical sensors.

Electrochemical detection method	Electrode	Sample	Antioxidant	Detection limit	Range linear	Ref
Cyclic Voltammetry	Glassy carbon	Spices	Curcumin	4.1×10^{-6} M	9.9×10^{-6} – 1.07×10^{-4} M	[53]
Amperometry	AgNP/Delph/GCE	Apple juice, lemon juice, peach juice, orange juice, green tea	Gallic acid	0.28 μmol/L	6×10^{-7} – 8.68×10^{-6} M	[58]
Differential Pulse Voltammetry	GCE	Human serum blood	Total antioxidant capacity	-	-	[59]
Differential Pulse Voltammetry	GCE/ZnAl-NO ₃ layered double hydroxide film	-	Gallic Acid	1.6 μM	4–600 μM	[44]
Differential Pulse Voltammetry, Cyclic voltammetry	GCE	-	Tertiary butyl hydroquinone	67 nM	1.0 μM–1.1 mM	[54]
Differential pulse polarography	Dropping mercury	-	Gallic acid	0.3 μM	1.0–50 μM	[60]
Differential pulse polarography	Ti ₃ Al0.5Cu0.5C ₂ /GCE	Kiwi	Rutin	0.015 μmol L ⁻¹	0.02–50.00 μmol/L	[61]
Square wave voltammetry	SPCE	-	• Ascorbic acid	• 0.09 mmol/I		[56]
Square wave voltammetry	SPCE	-	• N-acetylcysteine	• 0.04 mmol/I	-	[56]
Square wave voltammetry	SPCE	-	• Melatonin	• 0.07 mmol/I	-	[56]
Square wave voltammetry	4-[(4-decyloxyphenyl)-ethynyl]-1-methylpyridinium iodide modified glassy carbon	Mate herb extracts	Caffeic acid standard	9.0×10^{-7} M 8.7×10^{-6} M	9.9×10^{-7} M– 3.8×10^{-5} M 4.7×10^{-5} M– 9.9×10^{-5} M	[62]
Differential Pulse Voltammetry	(ZrO ₂ /Co ₃ O ₄ /rG)	Tea, juice and urine	Gallic acid and uric acid	2.5×10^{-8} M	2.2×10^{-7} – 5.5×10^{-5} M	[52]
Differential Pulse Voltammetry	Am-ZrO ₂ -CPE	Wine	Gallic acid	1.24×10^{-7} M	1×10^{-6} – 1×10^{-3} M	(46)
Differential Pulse Voltammetry	Nano-GO-SiO ₂ -nanoparticles-GCE	Red wine	Gallic acid	6.25×10^{-6} M	1×10^{-6} – 1×10^{-3} M	[63]

*- : Not Available.

4. Antioxidants-based Biosensors

Biosensors are a promising alternative method in several analytical processes in various fields. These techniques have been effectively utilized to detect various ailments, toxins, pathogens, and heightened levels of blood constituents, among other applications [64–66]. The biosensor comprises two primary components, specifically the bioreceptor and transducer elements. The bioreceptor component of the biosensor interacts interactively with the target analyte to ensure the sensor's selectivity. It happens because the identifier element in the biosensor takes advantage of specific biological interactions in compounds to increase the specific binding affinity for the desired molecule. Whereas transducers convert biological responses resulting from interactions with target analytes into measurable signals, it determines biosensor sensitivity [67]. The variety of transducers employed substantially affects the biosensor's sensitivity and selectivity.

The bioreceptors used in biosensors include enzymes [68], nucleic acids [69], antibodies [70], cells [71], and receptors. Sensitivity, selectivity, detection limit, stability, and linearity are used to evaluate the biosensor's performance. In comparison, the transducers used in biosensors are piezoelectric/gravimetric [72], optical [73], thermometric [74], and conductometric [75] (Figure 2).




Figure 2. Schematic of the general biosensor.

The biosensor used in this article uses an electrochemical biosensor taken from several literature studies. Electrochemical biosensors are starting to become a concern for development in measuring antioxidant activity because they can analyze biological samples, which can then be converted through biological processes and produce electrochemical signals. [16]. Biosensor indicators have high selectivity. The principle of the electrochemical biosensor is that when an analyte is at a different potential, it will cause oxidation to occur on the working electrode and then measure the signal from the movement of electrons. This signal can be measured with an electrochemical detector [76].

4.1. Enzyme-based biosensor

Developing enzyme-based biosensors as biocatalysts to accelerate chemical reactions for testing antioxidant activity needs attention to several aspects, such as immobilization and enzyme stability. In immobilizing the enzyme on the electrode, it is necessary to formulate the most efficient way because the immobilization of the enzyme affects the electron transfer rate. Therefore, enzyme immobilization must be integrated with nanomaterials or polymer membranes to produce good efficiency and stability and achieve the desired results. Enzyme-based electrochemical biosensors employ enzymes as bioreceptor elements, and the analysis of samples is predicated on inhibiting enzymatic activity. [77]. An enzymatic biosensor is a detection tool with the working principle of changing the measured substance concentration into a digital signal through a transducer. The components used to identify the biomolecules have suitable sensitive components, including an enzyme [78].

4.1.1. Tyrosinase

The enzyme tyrosinase is involved in the process of melanin biosynthesis, which contains copper and serves to catalyze the hydroxylation of mono-phenols to o-diphenols through the monophenolase reaction, as well as the oxidization of o-diphenols to o-quinone through the diphenolase reaction. *L-tyrosine* is a substrate that plays a role in monophenolase and diphenolase reactions. The enzymatic mechanism described involves a measurement approach involving the reversible electrochemical reduction of o-quinone, produced from phenol during enzymatic reactions. Each signal obtained from this measurement corresponds to the concentration of polyphenols in the solution. Enzyme-based biosensors are commonly utilized due to their remarkable sensitivity and selectivity. Enzymes involved in oxidation-reduction reactions have gained considerable interest in enzyme-based biosensors due to their remarkable ability to catalyze reactions dependent on electron transfer. The tyrosinase's enzymatic mechanism entails redox activity enabled by the reversible transfer of electrons between copper ions in their +1 and +2 oxidation states ($\text{Cu}^{+} \leftrightarrow \text{Cu}^{2+}$).

Electrochemical sensor development with the addition of nanoparticle-based materials whose goal is to increase sensor performance, namely biosensors based on tyrosinase/laccase enzymes with graphite-epoxy modified copper nanoparticles whose performance is reviewed in the determination of polyphenols, catechols, caffeine acids and catechins with cyclic voltammetry methods produces a linear range $> 200 \mu\text{M}$ [79].

Other studies are based on the tyrosinase enzyme with catechol as a substrate, where an excellent enzyme immobilization system produces reticulated BSA with chitosan for tyrosinase. The stability obtained by this tyrosinase-based biosensor lasts one to two days. Biosensor analysis using cyclic voltammetry (CV) and chronoamperometric (CA). The detection limit of each type of sensor is 0.5 μM for tyrosinase with a linear range of 1–340 μM [80].

An amperometric biosensor with the tyrosinase enzyme immobilized with glutaraldehyde on the SPCE evaluated catechins in black and green tea using cyclic voltammetry. This biosensor has a high sensitivity of 217 nA/ μM , a LOD of 0.03 μM , and 85% and 70% biosensor stability after 34 and 53 days. Then, the biosensor approach was compared with HPLC with recovery calculations of 90% and 96%, proving that tyrosinase as an enzyme-based biosensor can substitute for testing catechin derivatives in tea [81]. Furthermore, biosensors based on carbon paste electrodes were immobilized on Nafion films to measure hydroquinone and polyphenol antioxidants. The detection limit obtained was 1.6 $\mu\text{mol/L}$ with a repeatability of 1.2% RSD. Biosensors can determine the phenolic antioxidant capacity in plant and food samples without higher ascorbic acid concentrations [82].

The determination of chlorogenic acid is facilitated by a biosensor that employs the tyrosinase enzyme and screen-printed graphene-based electrodes modified with manganese phthalocyanine. The findings indicated that manganese phthalocyanine could enhance the activity of the tyrosinase enzyme and facilitate electron transfer. The linearity of the GPH-MnPc Tyr/SPE calibration curve against CGA, as determined through cyclic voltammetry and SWV, falls within the concentration range of 0.1–10.48 μM . The LOD and LOQ values are low and comparable to those of other tyrosinase-based biosensors in detecting phenolic compounds. The obtained biosensor results were subjected to comparison with the spectrophotometric method. The two methods yielded comparable validation outcomes, indicating that the biosensor approach for analyzing chlorogenic acid in nutraceutical products is characterized by high selectivity, rapid response, and holds the potential for monitoring phenolic compounds [83].

4.1.2. Laccase

Laccase, an enzyme categorized as benzenediol oxidoreductase and E.C. 1.10.3.2, is a multi-copper oxidase. Due to its ability to facilitate the four-electron reduction of dioxygen to water while simultaneously catalyzing the one-electron oxidation of four molecular sub-substrates, the catalyst in issue is considered green [84]. Laccase can aid in the oxidation of various organic and inorganic compounds, including n-methoxy-substituted phenols, polyphenols, ketones, ascorbates, phosphates, and diamines [85]. The enzyme has undergone thorough investigation due to its capacity

to catalyze the oxidation of diverse organic substrates and reduce molecular oxygen to water. Laccase is a novel enzyme that can catalyze the oxidation of diverse substrates without H_2O_2 within the reaction milieu. The electrode can facilitate the introduction of numerous analytes by utilizing two or more enzymes. The combination of laccase-tyrosinase enzymes is highly efficient in detecting phenolic compounds [86]. Laccases demonstrate responsiveness to ortho- and para-diphenol moieties, which include mono-, di-, and polyphenols, aminophenols, methoxy phenols, aromatic amines, and ascorbate, through reducing the four-electron oxygen to water (Figure 3).




Figure 3. Schematic of the oxidation reaction by laccase enzyme.

Laccase is a thermostable catalyst, so the biosensor performance increases more rapidly. The image above shows the sensor using three different screen-printed electrodes (SPE based on carbon nanotubes (CNT), gold nanoparticles (GNPs), and nanocarbon tubes and gold nanoparticles (CNT-GNP), which are then modified using laccase enzyme with the addition of glutaraldehyde as a crosslinking agent between the amino group in laccase and the aldehyde group of the reticulating agent, catechins were analyzed via CV and DPV techniques to explore their electrooxidation characteristics. The results showed that a detection limit established for catechins was 5.6×10^{-8} M on CNT-Lac/SPE, 1.3×10^{-7} M on GNP-Lac/SPE, and 4.9×10^{-8} M on CNT-GNP. The biosensor was exposed to the nutritional composition containing green tea to be analyzed for its catechin content using CNTGNP-Lac/SPE by DPV. So the relationship between CNT and GNP significantly increases the sensitivity and selectivity of the biosensor. Catechin content was assessed by paired t-test. Furthermore, the correlation between the response of CNT-GNP-Lac/SPE and the particular efficacy and antioxidant activity of nutrients, ascertained through traditional spectrophotometric techniques (DPPH, galvanoxyl, and ABTS), is examined about the swift development of biosensors for the assessment of comparative antioxidant activity [87].

Yang et al. [88] detected catechins using a laccase-based electrochemical sensor immobilized in a nanocomposite of 4-MBA gold nanoparticles and a polymer from CTS-g-N-CSIDZ. Biomaterials were employed to amplify the reactivity of electrochemical biosensors, which exhibit high specificity, rapid reactivity, economic viability, and uncomplicated configuration, obviating the necessity for arduous sample preparation or analytical instrumentation. The performance of laccase-based sensors as catechin sensors in industrial waste was evaluated through conventional methods, namely HPLC. The results obtained by the concentration of catechins in 3 industrial waste samples confirmed no significant disparity in the concentrations of the samples between the two methods. It shows that the electrochemical sensor can detect catechins. In conclusion, this sensor shows benefits such as rapid analysis duration, economic cost, and mobility.

In addition, measurements of electrochemical and electroanalytical antioxidants can be determined simultaneously. Laccase-based electronic detection and biosensors to measure the total phenolic content of honey samples from different countries showed a good correlation between antioxidant strength as measured by the electrochemical index and the FRAP and DPPH tests, as well as between TPC results from an approved biosensor and the Folin-Ciocalteau (FC) test [68].

4.1.3. Peroxidase-based biosensors

Peroxidase enzymes catalyze oxidation-reduction reactions through mechanisms involving free radicals [89]. Peroxidases are enzymes that facilitate the conversion of substrates into oxidized

products. HRP enzyme, or horseradish peroxidase, is frequently used in biochemistry and biotechnology. It is a reporting enzyme in affinity-based assays and biosensor recognition elements [90]. However, only some studies on peroxidase-based electrochemical biosensors in determining antioxidants still exist. Peroxidase-based electrochemical biosensors are starting to develop because they can transfer electrons at low working potentials. Consequently, there is a notable enhancement in the biosensor's selectivity [91]. The advantages possessed by other peroxide-based biosensors are good stability in wet test media and high sensitivity. HRP-based electrochemical biosensors have been extensively developed by immobilizing on classical working electrodes, such as carbon or precious metals. Enzyme-based biosensors have advanced in electrode immobilization and have many interface functions on the electrode surface [58].

A peroxidase-based electrochemical sensor method has been developed. It shows pretty good results based on the detection limit value and also the resulting linear range, including H_2O_2 measurements using the electrochemical detection method in the form of cyclic voltammetry using nanoparticle working electrodes encapsulated with HRP yielding LOD values above 0.01 μM and linear regression of 0.01–100 μM [92]. Measurement of L-lactate in wine and must samples using the electrochemical detection method in the form of amperometric voltammetry using a graphite rod working electrode combined with peroxidase-mimetic nanozymes produced LOD values above 2 μM and linear regression of 5 μM –14nm [91]. The peroxidase-based biosensor detects butylated hydroxyanisole (BHA) and propyl gallate (PG) in the food matrix. The HRP enzyme is immobilized by the spiny Au-Pt nanotube (SAP NTs) electrode, which has a wide surface area and is ideal for electron transfer. The SAP NTs structure was synthesized, and the intrinsic peroxidase was proved as an enzymatic biosensor. Linear sweep voltammetry (LSW) showed oxidation maxima for BHA and PG at 624 mV and 655 mV, respectively, with linearity ranges of 0.3–50 mg/L and 0.1–100 mg/L. BHA and PG were detectable at 0.046 and 0.024 mg/L (3s/slope), respectively. RSD experiments and recoveries were comparable to HPLC with good sensitivity, stability, and reducibility [93].

4.2. Cell/microorganism-based biosensors

Cell-based biosensors have gained significant popularity in various domains, such as drug screening, clinical and health, environmental testing, biomedicine, and national security [94–96]. Living cells are used as biometric components in electrochemical biosensors to convert biological impulses into electrical signals that can be quickly detected and measured [97]. Cell-based electrochemical biosensors are considered to be highly resilient biosensors that are capable of detecting biochemical effects through living cells. These biosensors possess several distinctive advantages, such as non-invasiveness, absence of labeling requirements, rapid response time, high efficiency, and versatile fabrication.

The hydrogen peroxide (H_2O_2) discharged from the cell reacts with a catalytic agent on the electrode's surface. This reaction facilitates reduction and oxidation reactions, subsequently leading to electrochemical signal alterations. The expeditious and precise discharge of H_2O_2 within cells is crucial in utilizing cell-based electrochemical biosensors for evaluating activity and screening antioxidant components. The study's methodology involves immobilizing cells onto a working electrode modified with manganese dioxide nanowires (MnO_2NWs) and gold nanoparticles (AuNP). It is achieved using a photosensitive hydrogel treated with carbon nanofibers (CNF) as a scaffolding material for three-dimensional cell culture. This study uses cells as receptor elements to investigate H_2O_2 release and assess the antioxidant potential of anthocyanins. The A549 cell-based electrochemical biosensor was utilized in the preparation process to detect extracellular H_2O_2 and evaluate the efficacy of the active constituents in terms of their antioxidant activity. The findings of this investigation indicate a strong association between the oxidation peak (I_p) and the concentration of H_2O_2 , as evidenced by the linear equation I_p (μA) = 58.199 CH_2O_2 +5.825 (where CH_2O_2 represents the concentration of H_2O_2). Moreover, the limit of detection for H_2O_2 was determined to be 0.02 μM , indicating a high level of sensitivity, reproducibility, and stability, as reported in the literature [71].

Ge et al. [98] also examined the antioxidant activity of a Chinese dried ham *Lactobacillus plantarum* extract. RAW264.7 cell-based electrochemical technique assessed this potential. RAW264

powers the cellular biosensor. Seven cells were placed in a hydrogel of sodium alginate and graphene oxide on a manganese dioxide surface changed with a gold electrode. To make a 3D hydrogel scaffold, cells/NaAlg/GO/a-MnO₂/GE was immersed in calcium. After PMA, RAW264.7 cells produced H₂O₂. RAW264 cells released H₂O₂. The MnO₂/GE absorbs seven cells and catalyzes them at its active site. MnO₂ reduces and electro-oxidizes on the electrode surface, increasing the oxidation-reduction probe and electron transfer rate. The biosensor approach was used to detect H₂O₂ from RAW264.7 macrophage cells quickly. The study has a detection limit of 0.02 μM and a linear response from 0.05 to 0.85. At 1010 CFU/mL, *L. plantarum* had the highest relative antioxidant capacity (RAC) of 88.94%. Thus, this study uses RAW264.7 cells to assess *L. plantarum*'s antioxidant potential quickly, sensitively, and quantitatively via electrochemical biosensing. This approach quickly screens specimens' antioxidant properties.

Microorganisms producing superoxide dismutase enzymes (*D. radiodurans*) as baroreceptors were immobilized in zeolite nanocomposites using electrochemical measurements, namely cyclic voltammetry. Xanthine oxidase (XO) catalyzes the enzymatic process of xanthine as a substrate, generating superoxide radicals. The presence of zeolite has the potential to augment the response, concomitantly elevating the value of Km. It is because the immobilization of SOD in zeolite for 24 hours does not necessarily imply a lower concentration of SOD involved in the reaction with superoxide radicals compared to prior research [99]. The detection of uric acid as a non-enzyme antioxidant using *Lactobacillus plantarum* immobilized using zeolite reported the results of electrode stability until day 18 [100]. Furthermore, the detection of antioxidants using superoxide dismutase-producing microorganisms from *D. radiodurans* bacteria as receptors immobilized on the surface of SPCE using electrochemical measurements, namely cyclic voltammetry. The results showed that the stable *D. radiodurans* biofilm on the SPCE surface remained stable for 35 days. Optimum SPCE biofilms made have good stability with less than 5% RSD. Repeatability measurements indicated that the optimum SPCE biofilm had fairly good repeatability. The voltammogram produced with three different SPCE biofilms is identical and has an oxidation peak of 0.750 V [101].

4.3. DNA-based-biosensors

DNA-based biosensors have advantages, including superior biocompatibility, good thermal stability, alternative functionalization, and detecting specific targets [102–106]. Antioxidants in food samples can be detected using DNA as a bioreceptor in a biosensor. The DNA emulate the mechanisms of interaction with analytical agents that take place within the human body by immobilizing DNA on the transducer surface can be achieved through the utilization of genetic material as a biological recipient [107]. DNA has been identified as a highly promising bioreceptor for biosensing applications owing to its extended biological activity, exceptional addressability, and tunable stiffness. DNA-based biosensors have been widely developed and show promising results, such as aptamer, which has better thermal stability [108]. Adaptable biological affinity [109–111] and improved enzyme attack resistance [112]. DNA can also be used to construct supermolecular structures that can be programmed as templates to realize precise positioning and control spatial modifications, significantly improving biosensor performance [113]. DNA-based biosensors encompass various types such as functional DNA, DNA hybridization, and DNA template-based biosensors.

Typically, electrochemical biosensors that rely on DNA as the bioreceptor element utilize immobilized DNA molecules onto the working electrode surface to facilitate the detection of DNA-analyte interactions. Moreover, these interactions elicit modifications in the structure of DNA and its electrochemical characteristics, converting a stimulus into an electrochemical signal [114]. DNA-based biosensors can be used to determine the antioxidant capacity of samples on a routine basis because the changes in the base oxidation peaks are highly distinguishable before and after interactions with analytes, so there is no need for labeling or amplification strategies that can reduce analysis time and complexity [16].

Then, immobilization of purine bases (adenine and guanine) onto the surface of carbon glass electrodes (GCE) utilizes hydroxyl radicals for purine base degradation), utilizing hydroxyl radicals

for purine base degradation. Ascorbic acid, caffeic acid, coumaric acid, and resveratrol are compounds with antioxidant properties that can eliminate hydroxyl radicals and safeguard adenine and guanine located on the surface of GCE. The study assessed the interactions between immobilized purine bases and free radicals, both in the presence and absence of antioxidants, by analyzing alterations in anodic peak currents utilizing square wave voltammetry (SWV). The outcomes derived from the five antioxidants exhibited varied efficiencies, from 47% to 79%, in their capacity as hydroxyl radical scavengers. The antioxidant standard with the highest sensitivity, ascorbic acid, yielded the greatest TAC value. The utilization of purine-base-based biosensors can enhance beverages' overall antioxidant capacity. The expeditious and uncomplicated nature of the analysis is augmented by its compatibility with portable instrumentation, as noted in reference [115].

Peng et al. [116] used guanine, immobilized on the electrode modified by adding MoS₂ nanosheet. This results in an electrochemical biosensor with a large specific surface area and high electrocatalytic capability. The development of a guanine-based biosensor modified by nanosheets used in measuring the antioxidant capacity of three types of flavonoids, namely quercetin, fisetin, and catechins, compared to ascorbic acid, shows many advantages, including an extensive linear range, low detection limit, and good stretching. Quercetin, fisetin, and catechin were measured and found to be 45.82%, 34.39%, and 16.99%, respectively. Ascorbic acid is a comparator antioxidant; its capacity value was determined to be 51.84%. [117].

The determination of chlorogenic acid (CGA) in coffee samples using the cyclic voltammetry method and SPCEs electrodes modified with single-walled carbon nanotubes (SWCNT-COOH) has been reported. dsDNA interacts with OH free radicals to measure antioxidant activity in a DNA-based biosensor [118]. The current study used square wave voltammetry and deoxy-adenylic acid oligonucleotide (dA20) immobilized on carbon paste electrodes to identify polyphenols in green tea, black tea, peppermint, and senna samples. According to reference 29, the procedure yielded a linear range of 1.0 to 8.0 mg L⁻¹ and a LOD of 0.059 mg L⁻¹. Glutathione and ascorbic acid concentrations were measured using pulse differential voltammetry. This study used pencil graphite electrodes (PGE) modified with MWCNTs and chitosan (CHIT)/ds-DNA. Reference describes this procedure [119].

5. Nanosensors

Electrochemical techniques are becoming faster, more accurate, and less expensive to measure antioxidant capacity. Traditional spectrophotometric methods have problems, such as long analysis and sample preparation times, reagents that could be better for the environment, expensive reagents, and unknown reaction times [13].

Electroanalytical methods use transfer studies to exclude it. Electrochemical principles explain electron transfer and antioxidant oxidation reactions, although visual approaches are important. Electron transfer methods use a simple oxidant-redox process to generate signals. Electrochemical methods are appropriate for studying electron transfer. Due to their great sensitivity, voltammetric methods are suitable for qualitative analysis and quantifying trace phenolic compounds and antioxidants. Due to their electron-donating capacity, voltammetric methods can detect antioxidants like polyphenols at low potentials [120]. Several literature studies on nanosensors using electrochemical methods can be seen in Table 2.

Electrochemical sensing is the most prevalent analytical method for identifying analytes due to its speed, sensitivity, accuracy, and affordability. Electrochemical sensors are useful for directly assessing antioxidants and detecting dietary antioxidant capacity due to these qualities. Active electrode materials have been used to make antioxidant biosensors and electrochemical sensors [13].

Manoranjitham et al. used a BHA electrochemical nanosensor to detect synthetic antioxidants. Electro-polymerization created a POC/MWCNTs electrode sensor employing o-cresol phthalein complex-one (OC). The sensor's detection limit was 0.11 mM [121]. In another study, an electrochemical sensor detected BHA in food and other items. Modifying multi-layer graphite paper with AuNPs and NiO nanoparticles created the sensor. AuNPs' conductivity complemented NiO

nanoparticles' electrocatalytic activity. EGP's broad surface area boosts linear range, lowering LOD and increasing sensitivity. The NiO/AuNPs/EGP combination identified samples well [122].

Ziyatdinova devised an exact and discerning sensor capable of detecting TBHQ and BHA concurrently. The sensitive layer was formed by grouping MWNT with electropolymerized carminic acid. The augmentation of the electrode surface area and electron transfer rate resulted in a notable enhancement in the response of TBHQ and BHA. In turn, facilitated the observation of their oxidation peaks when they coexisted. The utilization of Poly(carminic acid) as a surface modifier for electrodes is being reported for the first time. The sensor that was obtained exhibited a high level of sensitivity and could identify template molecules. The sensor's response exhibits linearity within the 0.50-75 μM range for TBHQ and 0.25-75 μM range for BHA, and the respective limits of detection are 0.36 μM and 0.23 μM [123].

Motia et al. developed an electrochemical sensor utilizing molecularly imprinted polymer (MIP) and benzenehexacarboxylic acid (BHA). The electrochemical sensor that underwent chitosan modification and was coated with gold nanoparticles and molecularly imprinted polymers demonstrated high efficacy in detecting BHA. The findings indicate that the electrochemical sensor exhibits remarkable selectivity and robust sensitivity. The created sensor exhibited a limit of detection (LOD) of 0.001 $\mu\text{g/mL}$. In a recent study, three commercially available food items, namely chewing gum, mayonnaise, and potato chips, were utilized as the original testing medium for evaluating the performance of the molecularly imprinted polymer (MIP) sensor. The findings of the study indicated favorable outcomes [124]. Balram et al. synthesized a novel nanocomposite of CuO nanofibers and NH₂-functionalized carbon nanotubes, which exhibited superior electrocatalytic activity for detecting cytotoxic TBHQ. The electrocatalytic performance of CuO nanofibers (NFs) supported on amino-functionalized carbon nanotubes (NH₂-CNTs) on SPCE was investigated using cyclic voltammetry (CV) and differential pulse voltammetry (DPV) techniques. The sensor exhibited a high level of sensitivity, measuring at 37.7 $\mu\text{A/M/cm}^2$, a substantial detection limit of 3 nM, and a broad linear range, as reported in reference [125]. Yan et al. proposed an alternative approach for the detection of TBHQ utilizing an electrochemical sensor. VMSF/ErGO has modified the recognition layer of GCE film for utilization. The sensitivity of the VMSF/ErGO/GCE sensor towards TBHQ can be attributed to the preconcentration effect of hydrogen bonds in VMSF and the elevated electroactivity of ErGO, as reported in reference [126].

Table 2. Tabulated some studies about electrochemical nanosensor research for analyzing endogenous antioxidants.

Method	Electrode	Medium	Antioxidants	Matrix	LOD	Linear Range	Ref.
DPV	MIM-PACO/GCE	0.25 M ABS (pH 6.5)	Curcumin	Turmeric extract	5.0 nM	10 nM–2.0 μM	[127]
DPV	SNO NRs/GCE	0.1 M PBS (pH 5.0)	Quercetin	Apple and grape juice	1.98 nM	0.01–68.53 μM	[128]
DPV	MMIP	0.1 M PBS (pH 1.0)	Rosmarinic acid	<i>Salvia officinalis</i> , <i>Zataria multiflora</i> , <i>Mentha longifolia</i> , and <i>Rosmarinus officinalis</i>	0.085 μM	0.1–100 μM 100–500 μM	[129]
DPV	EGDMA-MIP/IL-GR/GCE	0.04 M BRBS (pH 2.0)	Rutin	Tablet	0.12 μM	0.3–1 μM	[130]
DPV	GCE/rGO/ZIF-8/MIP		Rutin	Tablet and orange juice	0.0001 μM	0.05–100 μM 0.0005 - 0.05 μM	[131]
DPV	MIP/AuNPs/EGP	0.1 M PBS (pH 5.0)	TBHQ	Edible oil	0.07 μM	0.08–100 μM	[122]
DPV	MIP/CHIT + AuNPs/SPCE	PBS (pH 7.0)	BHA	Chewing gum, mayonnaise and potato chips	0.001 $\mu\text{g/mL}$	0.01–20 $\mu\text{g/mL}$	[124]
DPV	CuO.NFs/NH ₂ -CNTs/SPCE	0.05 M PBS (pH 7.0)	TBHQ	Coconut oil, sesame oil, soybean oil	3 nM	0.013.9 μM 3.9–147.6 μM	[125]
DPV	VMSF/ErGO/GCE		TBHQ		0.23 nM	0.001–0.5 μM	[126]

		0.1 M PBS (pH 4.0)	Edible oil, Toning lotion		0.5–120 μ M		
			BHA	Potato chips	0.11 μ M	0.33–110 μ M	[121]
AMP	poly O-cresolphthalein/MWCNT electrode	0.1 M PBS (pH 7.0)	BHA	Potato chips	0.11 μ M	0.33–110 μ M	[121]
AMP	poly(carminic acid)/MWNT/GCE	BRBS (pH 2.0)	BHA TBHQ	Linseed oil	0.23 μ M 0.36 μ M	0.25–75 μ M 0.50–75 μ M	[123]
AMP	4-aminobenzoic acid/Toray carbon fiber electrode	0.1 M PBS (pH 7.0)	Bilirubin	Serum	15 μ M	150–890 μ M	[132]
AMP	AuNPs/RGO/SPCE	100 mM PBS (pH 3.5)	Vitamin C	Commercial, pasteurized, and skimmed cow's milk	0.088 μ g/mL	50–500 μ M	[133]
AMP	Mesoporous CuCo ₂ O ₄ /GCE	0.15 M NaOH solution	Vitamin C	Vitamin C tablets Effervescent tablets	0.21 μ M	1–100 μ M	[134]
CV	H-BDDP-printed electrode	1/15 M PBS (pH 7.0)	L-Cysteine	Bovine plasma	0.620 μ M	1–194 μ M	[135]
CV	GOCuNP/CPE	1.0 M KCl (pH 7.0)	N-acetylcysteine	-	2.97×10^{-5} M	3.0×10^{-4} – 6.0×10^{-3} M	[136]

*- : Not Available.

6. Biosensors Trends and Perspective

In recent years, there has been a significant expansion in the field of biosensors. However, more biosensors need to be developed specifically to address issues related to sustainable agriculture. Recent publications have primarily focused on utilizing established sensor principles for detecting food products and other related applications. From a positive perspective, contemporary situations could be encompassed by commercially available sensor technology, thereby emphasizing the importance of advancing integrated application development. The research group focused on leadership status has disclosed the creation of an electrical stimulation machine that has the potential to enhance agricultural productivity and provide autonomous, weather-resistant sensing services. The biosensors have encountered challenges with anti-interference measures, self-calibration techniques, and prolonged monitoring periods, necessitating further attention.

The most important improvements in enzyme-based biosensors involve how the biological materials on the electrode surface are held in place and how the interface works. Even though this technology works well and is important for both applied and basic science, four important things need to be thought about before enzyme biosensors can be used commercially to track active compounds and their antioxidant power:

- Stability and immobilization

In making biosensors, it is hard to immobilize enzymes in a way that works well and speeds up the rate at which electrons are transferred. Nanomaterials and polymers have the potential to serve as carriers or hosts for immobilized enzymes, thereby facilitating electron transfer, enhancing the stability of biosensors, and prolonging their lifespan.

- Several kinds of enzymes

Due to the high specificity of enzyme reactions, it is not feasible for a single type of enzyme to effectively locate all antioxidants or assess the antioxidant properties of all active substances. The enzyme Laccase is incapable of degrading monophenols such as 3-amino phenol due to the positioning of the amino group in the meta position. Consequently, the biosensor necessitates greater specificity, challenging certain substrates such as monophenols. Developing diverse enzyme-based biosensors capable of detecting specific antioxidants through enzymatic mechanisms would be an effective avenue for exploration.

- Interference with the matrix

The issue of matrix interference poses a significant challenge for numerous research methodologies, including biosensor techniques. It is due to the complex nature of real samples. For

the matrix interference to be kept to a minimum, not only do new ways of pre-treating samples need to be found, but the specificity and selectivity of the biosensor also need to be improved.

- Sensitivity and usage

Enzymes need to be attached to the electrode surface for biosensors' effectiveness through bioconjugation. High-specificity sensing is necessary to engineer biocompatible materials that satisfy the criteria for achieving heightened sensitivity. The significance of engaging with the matrix lies in its ability to reduce the effective concentration of enzymes. The forthcoming advancement of enzymatic biosensors will prioritize enhancing the immobilization technology and modifying the biological enzyme to augment the biosensor's utility.

Author Contributions: For review articles with several authors, all authors were involved in conceptualization, writing review and editing, supervision, and funding acquisition. All authors have read and agreed to the published version of the manuscript.

Funding: This research received no external funding.

Institutional Review Board Statement: Not applicable.

Informed Consent Statement: Not applicable.

Conflicts of Interest: The authors declare no conflict of interest.

References

1. Li, Q.; Guan, X.; Wu P.; Wang, X.; Zhou, L.; Tong, Y. Early Transmission Dynamics in Wuhan, China, of Novel Coronavirus-Infected Pneumonia. *N Engl J Med.* **2020**, *382*, 1199–207.
2. Lushchak, V.I. Free radicals, reactive oxygen species, oxidative stress and its classification. *Chem Biol Interact.* **2014**, *5*, 164-75. <https://doi.org/10.1016/j.cbi.2014.10.016>.
3. Senoner, T.; Dichtl, W. Oxidative Stress in cardiovascular diseases: still a therapeutic target. *Nutrients.* **2019**, *11*, 2090.
4. Prasad, M. Sedlarova, A. Balukova, M. Rac, P. Pospisil. Reactive Oxygen Species as a Response to Wounding: In Vivo Imaging in *Arabidopsis thaliana*. *Front Plant Sci.* **2020**, *10*, 1-10. <https://doi.org/10.3389/fpls.2019.01660>
5. Zhang, L.; Liu, Y. Potential interventions for novel coronavirus in China: A systematic review. *J Med Virol.* **2020**, *92*, 479–90
6. Alkadi, H. A review on free radicals and antioxidants. *Infect Disord Drug Targets.* **2020**, *20*, 16-26. <https://doi.org/10.2174/1871526518666180628124323>.
7. Marrocco I, Altieri F, Peluso I. Measurement and Clinical Significance of Biomarkers of Oxidative Stress in Humans. *Oxid Med Cell Longev.* **2017**, *2017*, 1-32. <https://doi.org/10.1155/2017/650104>
8. Pisoschi, A.M.; Pop, A. The role of antioxidants in the chemistry of oxidative: A review. *European J Med Chem.* **2015**, *97*, 55-74.
9. Sandrasari, D.A.; Andarwulan, N.; Faridah, D.N.; Dewi, F.N.A. Indonesian indigenous plants as a source of antioxidants to treat gastrointestinal disorders. *Food Res.* **2021**, *5*, 195–204.
10. Costa, C.; Tsatsakis, A.; Mamoulakis, C.; Teodoro, M.; Briguglio, G.; Caruso, E.; Tsoukalas, D.; Margina, D.; Dardiotis, E.; Kouretas, D.J.F.R. Current evidence on the effect of dietary polyphenols intake on chronic diseases. *Food Chem Toxicol.* **2017**, *110*, 286–299. <https://doi.org/10.1016/j.fct.2017.10.023>
11. Sharma, R.; Padwad, Y.J. Perspectives of the potential implications of polyphenols in influencing the interrelationship between oxi-inflammatory stress, cellular senescence and immunosenescence during aging. *Trends Food Sci & Technol.* **2020**, *98*, 41-52.
12. Iswantini, D.; Yulian, M.; Mulijani, S.; Trivadila. Inhibition kinetics of *Sida rhombifolia* L. extract toward xanthine oxidase electrochemical method. *Indo. J. Chem..* **2014**, *14*, 71-77.
13. Pisoschi, A.M.; Pop, A.; Iordache, F.; Stanca, L.; Predoi, G.; Serban, A.M. Oxidative stress mitigation by antioxidants an overview on their chemistry and influences on health status. *Eur J Med Chem.* **2021**, *112891*. <https://doi.org/10.1016/j.ejmech.2020.112891>.
14. Ruskovska, T.; Budić-Leto, I.; Corral-Jara, K.F.; Ajdžanović, V.; Arola-Arnal, A.; Bravo, F.I.; Deligiannidou, G.E.; Havlik, J.; Janeva, M.; Kistanova, E.; Kontogiorgis, C.; Krga, I.; Massaro, M.; Miler, M.; Milosevic, V.;

Morand, C.; Scoditi, E.; Suarez, M.; Vauzour, D.; Milenkovic, D. Systematic bioinformatic analyses of nutrigenomic modifications by polyphenols associated with cardiometabolic health in humans—Evidence from targeted nutrigenomic studies. *Nutrients*. **2021**, *13*, 1–28. <https://doi.org/10.3390/nu13072326>

15. Butterfield, D.A.; Halliwell, B. Oxidative stress, dysfunctional glucose metabolism and Alzheimer disease. *Nat Rev Neurosci.* **2019**, *20*, 148–160. <https://doi.org/10.1038/s41583-019-0132-6>

16. Ye, Y.; Ji, J.; Sun, Z.; Shen, P.; Sun, X. Recent advances in electrochemical biosensors for antioxidant analysis in foodstuff. *TrAC - Trends Anal Chem.* **2020**, *122*, 115718. doi:10.1016/j.trac.2019.115718.

17. Srinivasan, K. Antioxidant Potential of Spices and Their Active Constituents. *Crit Rev Food Sci Nutr.* **2014**, *54*, 352–372. doi:10.1080/10408398.2011.585525.

18. Yashin, A.; Yashin, Y.; Xia, X.; Nemzer, B. Antioxidant activity of spices and their impact on human health: A review. *Antioxidants*. **2017**, *6*, 1–18. doi:10.3390/antiox6030070.

19. Nimse, S.; Pal, D. Free radicals, natural antioxidants, and their reaction mechanisms. *RSC Adv.* **2015**, *5*, 27986–28006

20. Ojha, K.; Dubey, S.; Chandrakar, J.; Minj, R.A.; Dehariya, R.; Dixit, A.K. A review on different methods of determination of antioxidant activity assay of herbal plant. *J of Sci Bio Pharm and chem sci.* **2018**, *4*, 707–732. <https://doi.org/10.26479/2018.040656>

21. Chen, J.; Liu, Y.; Zhao, Z.; Qiu, J. Oxidative stress in the skin: Impact and related protection. *Int J Cosmet Sci.* **2021**, *43*, 495–509. doi:10.1111/ics.12728.

22. Działo, M.; Mierziak, J.; Korzun, U.; Preisner, M.; Szopa, J.; Kulma, A. The potential of plant phenolics in prevention and therapy of skin disorders. *Int J Mol Sci.* **2016**, *17*, 1–41. doi:10.3390/ijms17020160

23. Kobus-Cisowska, J.; Flaczyk, E.; Rudzińska, M.; Kmiecik, D. Antioxidant properties of extracts from Ginkgo biloba leaves in meatballs. *Meat Sci.* **2014**, *97*, 174–180. doi:10.1016/j.meatsci.2014.01.011

24. Rolim, W.R.; Pelegrino, M.T.; de Araújo Lima, B.; Ferraz, L.S.; Costa, F.N.; Bernardes, J.S.; Rodrigues, T.; Brocchi, M.; Seabra, A.B. Green tea extract mediated biogenic synthesis of silver nanoparticles: Characterization, cytotoxicity evaluation and antibacterial activity. *Appl Surf Sci.* **2019**, *463*, 66–74. doi:10.1016/j.apsusc.2018.08.203

25. Vinci, G.; D'ascenzo, F.; Maddaloni, L.; Prencipe, S.A.; Tiradritti, M. The Influence of Green and Black Tea Infusion Parameters on Total Polyphenol Content and Antioxidant Activity by ABTS and DPPH Assays. *Beverages*. **2022**, *8*, doi:10.3390/beverages8020018.

26. Thongsuk, P.; Sameenoi, Y. Colorimetric determination of radical scavenging activity of antioxidants using Fe3O4 magnetic nanoparticles. *Arab J Chem.* **2022**, *15*, 103475. doi:10.1016/j.arabjc.2021.103475.

27. Wang, X.; Luo, Y.; Huang, K.; Cheng, N. Advanced Agrochem Biosensor for agriculture and food safety : Recent advances and future perspectives. *Adv Agrochem.* **2022**, *1*, 3–6. doi:10.1016/j.aac.2022.08.002.

28. Munteanu, I.G.; Apetrei, C. A Review on Electrochemical Sensors and Biosensors Used in Chlorogenic Acid Electroanalysis. *Int J Mol Sci.* **2021**, *22*, 13138.

29. Barroso, M.F.; Ramalhosa, M.J.; Alves, R.C.; Dias, A.; Soares, C.M.D.; Oliva-Teles, M.T.; Delerue-Matos, C. Total antioxidant capacity of plant infusions: Assessment using electrochemical DNA-based biosensor and spectrophotometric methods. *Food Control*. **2016**, *68*, 153–161. doi:10.1016/j.foodcont.2016.03.029

30. Karimi-Maleh, H.; Orooji, Y.; Karimi, F.; Alizadeh, M.; Baghayeri, M.; Rouhi, J.; Tajik, S.; Beitollahi, H.; Agarwal, S.; Gupta, V.K.; Rajendran, S.; Ayati, A.; Fu,L.; Sanati, A.L.; Tanhaei, B.; Sen, F.; Shabani-nooshabadi, M.; Asrami, P.N.; Al-Othman, A. A critical review on the use of potentiometric based biosensors for biomarkers detection. *Biosens Bioelectron.* **2021**, *184*, 113252. doi:10.1016/j.bios.2021.113252.

31. Ye, Y.; Sun, X.; Zhang, Y.; Han, X.; Sun, X. 2022. A novel cell-based electrochemical biosensor based on MnO2 catalysis for antioxidant activity evaluation of anthocyanins. *Biosens Bioelectron.* **2021**, *202*, 113990. doi:10.1016/j.bios.2022.113990

32. Romero, K.J.; Galliher, M.S.; Raycroft, M.A.R.; Chauvin, J.P.R.; Bosque, I.; Pratt, D.A.; Stephenson, C.R.J. Electrochemical Dimerization of Phenylpropenoids and the Surprising Antioxidant Activity of the Resultant Quinone Methide Dimers. *Angew Chemie - Int Ed.* **2018**, *57*, 17125–17129. doi:10.1002/anie.201810870.

33. Ricci, A.; Parpinello, G.P.; Teslić, N.; Kilmartin, P.A.; Versari, A. Suitability of the cyclic voltammetry measurements and DPPH• spectrophotometric assay to determine the antioxidant capacity of food-grade oenological tannins. *Molecules*. **2019**, *24*, 1–12. doi:10.3390/molecules24162925.

34. Petković, B.B.; Stanković, D.; Milčić, M.; Sovilj, S.P.; Manojlović, D. Dinuclear copper(II) octaazamacrocyclic complex in a PVC coated GCE and graphite as a voltammetric sensor for determination of gallic acid and antioxidant capacity of wine samples. *Talanta*. **2015**, *132*, 513–519. doi:10.1016/j.talanta.2014.09.025.

35. Rajput, V.D.; Harish.; Singh, R.K.; Verma, K.K.; Sharma, L.; Quiroz-Figueroa, F.R.; Meena, M.; Gour, V.S.; Minkina, T.; Sushkova, S.; Mandzhieva, S. Recent developments in enzymatic antioxidant defence mechanism in plants with special reference to abiotic stress. *Biology (Basel)*. **2021**, *10*, 267. doi:10.3390/biology10040267

36. Losada-Barreiro, S.; Bravo-Díaz, C. Free radicals and polyphenols: The redox chemistry of neurodegenerative diseases. *Eur J Med Chem*. **2017**, *133*, 379–402. doi:10.1016/j.ejmech.2017.03.061

37. Vo, T.T.T.; Chu, P.M.; Tuan, V.P.; Te, J.S.L.; Lee, I.T. The promising role of antioxidant phytochemicals in the prevention and treatment of periodontal disease via the inhibition of oxidative stress pathways: Updated insights. *Antioxidants*. **2020**, *9*, 1–28. doi:10.3390/antiox9121211.

38. Gupta, D. Methods for determination of antioxidant capacity: A review. *Int J Pharm Sci Res*. **2015**, *6*, 546–566. doi:10.13040/IJPSR.0975-8232.6(2).546-66. Martínez, A.; Galano, A.; Vargas, R. Free radical scavenger properties of α -mangostin: Thermodynamics and kinetics of HAT and RAF mechanisms. *J Phys Chem B*. **2011**, *115*, 12591–12598. doi:10.1021/jp205496u.

39. Martinez-Perez, C.; Ward, C.; Cook, G.; Mullen, P.; McPhail, D.; Harrison, D.J.; Langdon, S.P. Novel flavonoids as anti-cancer agents: Mechanisms of action and promise for their potential application in breast cancer. *Biochem Soc Trans*. **2014**, *42*, 1017–1023. doi:10.1042/BST20140073.

40. Martinez-Perez, C.; Ward, C.; Cook, G.; Mullen, P.; McPhail, D.; Harrison, D.J.; Langdon, S.P. Novel flavonoids as anti-cancer agents: Mechanisms of action and promise for their potential application in breast cancer. *Biochem Soc Trans*. **2014**, *42*, 1017–1023. doi:10.1042/BST20140073.

41. Awouafack, Maurice D., et al. 'Isolation and Structure Characterization of Flavonoids'. Flavonoids - From Biosynthesis to Human Health, InTech, 23 Aug. 2017. doi: 10.5772/67881.

42. Tiwari, S.C.; Husain, N. Biological Activities and Role of Flavonoids in Human Health-A Review. *Indian J.Sci.Res*. **2017**, *12*, 193-196.

43. Badhani, B.; Sharma, N.; Kakkar, R. Gallic acid: A versatile antioxidant with promising therapeutic and industrial applications. *RSC Adv*. **2015**, *5*, 27540–27557. doi:10.1039/c5ra01911g.

44. Kahl, M.; Golden, T.D. Electrochemical determination of phenolic acids at a Zn/Al layered double hydroxide film modified glassy carbon electrode. *Electroanalysis*. **2014**, *26*, 1664–1670. doi:10.1002/elan.201400156

45. Carocho, M.; Ferreira, I.C.F.R. A review on antioxidants, prooxidants and related controversy: Natural and synthetic compounds, screening and analysis methodologies and future perspectives. *Food Chem Toxicol*. **2013**, *51*, 15–25. doi:10.1016/j.fct.2012.09.021

46. Chikere, C.O.; Faisal, N.H.; Kong-Thoo-Lin, P.; Fernandez, C. Interaction between amorphous zirconia nanoparticles and graphite: Electrochemical applications for gallic acid sensing using carbon paste electrodes in wine. *Nanomaterials*. **2020**, *10*, 1-25 doi:10.3390/nano10030537.

47. Zheng, Y.; Karimi-maleh, H.; Fu, L. Evaluation of Antioxidants Using Electrochemical Sensors: A Bibliometric Analysis. *Sensors*. **2022**, *22*, 1–25. doi:10.3390/s22093238.

48. Nejad, F.G.; Tajik, S.; Beitollahi, H.; Sheikhshoiae, I. Magnetic nanomaterials based electrochemical (bio)sensors for food analysis. *Talanta*. **2021**, *228*, 122075. doi:10.1016/j.talanta.2020.122075.

49. Ziyatdinova, G.; Guss, E.; Yakupova, E. Electrochemical sensors based on the electropolymerized natural phenolic antioxidants and their analytical application. *Sensors*. **2021**, *21*, 1-38. doi:10.3390/s21248385.

50. Gajdár, J.; Gonč, T.; Jampílek, J.; Brázdrová, M.; Bábková, Z.; Fojta, M.; Barek, J.; Fischer, J. Voltammetry of a Novel Antimycobacterial Agent 1-Hydroxy-N-(4-nitrophenyl)naphthalene-2-carboxamide in a Single Drop of a Solution. *Electroanalysis*. **2018**, *30*, 38–47. doi:10.1002/elan.201700547.

51. Haque, M.A.; Morozova, K.; Ferrentino, G.; Scampicchio, M. Electrochemical Methods to Evaluate the Antioxidant Activity and Capacity of Foods: A Review. *Electroanalysis*. **2021**, *33*, 1419–1435. doi:10.1002/elan.202060600

52. Shahamirifard, S.A.; Ghaedi, M.; Razmi, Z.; Hajati, S. A simple ultrasensitive electrochemical sensor for simultaneous determination of gallic acid and uric acid in human urine and fruit juices based on zirconia-choline chloride-gold nanoparticles-modified carbon paste electrode. *Biosens Bioelectron*. **2018**, *114*, 30–36. doi:10.1016/j.bios.2018.05.009.

53. Ziyatdinova, G.K.; Nizamova, A.M.; Budnikov, H.C. Voltammetric determination of curcumin in spices. *J Anal Chem.* **2012**, *67*, 591–594. doi:10.1134/S1061934812040132.
54. Wang, X.; Jiao, C.; Yu, Z. Electrochemical biosensor for assessment of the total antioxidant capacity of orange juice beverage based on the immobilizing DNA on a poly l-glutamic acid doped silver hybridized membrane. *Sensors Actuators, B Chem.* **2014**, *192*, 628–633. doi:10.1016/j.snb.2013.11.025.
55. Souza, L.P.; Calegari, F.; Zarbin, A.J.G.; Marcolino-Júnior, L.H.; Bergamini, M.F. Voltammetric determination of the antioxidant capacity in wine samples using a carbon nanotube modified electrode. *J Agric. Food. Chem.* **2011**, *59*, 7620–7625. doi:10.1021/jf2005589.
56. Kračmarová, A.; Pohanka, M. Elektrochemické stanovení nízkomolekulárních antioxidantů v séru. *Chem List.* **2014**, *108*, 64–69.
57. Munteanu, I.G.; Apetrei, C. Electrochemical determination of chlorogenic acid in nutraceuticals using voltammetric sensors based on screen-printed carbon electrode modified with graphene and gold nanoparticles. *Int. J. Mol. Sci.* **2021**, *22*, 8897. doi:10.3390/ijms22168897.
58. Ghaani, M.; Nasirizadeh, N.; Ardakani, S.A.Y.; Mehrjardi, F.Z.; Scampicchio, M.; Farris, S. Development of an electrochemical nanosensor for the determination of gallic acid in food. *Anal Methods.* **2016**, *8*, 1103–1110. doi:10.1039/c5ay02747k.
59. Korotkova, E.I.; Freinbichler, W.; Linert, W.; Dorozhko, E.V.; Bukkel, M.V.; Plotnikov, E.V.; Voronova, O.A. Study of total antioxidant activity of human serum blood in the pathology of alcoholism. *Molecules.* **2013**, *18*, 1811–1818. doi:10.3390/molecules18021811
60. Yilmaz, Ü.T.; Kekillioglu, A.; Mert, R. Determination of Gallic acid by differential pulse polarography: Application to fruit juices. *J Anal Chem.* **2013**, *68*, 1064–1069. doi:10.1134/S1061934813120113.
61. Senocak, A.; Sanko, V.; Tumay, S.O.; Orooji, Y.; Demirbas, E.; Yoon, Y.; Khataee, A. Ultrasensitive electrochemical sensor for detection of rutin antioxidant by layered Ti3Al0.5Cu0.5C2 MAX phase. *Food Chem. Toxicol.* **2022**, *164*, 1-9. <https://doi.org/10.1016/j.fct.2022.113016>
62. Silva, T.R.; Westphal, E.; Gallardo, H.; Vieira, I.C. Ionic organic film sensor for determination of phenolic compounds. *Electroanalysis.* **2014**, *26*, 1801–1809. doi:10.1002/elan.201400197.
63. Chikere, C.; Faisal, N.H.; Lin, P.K.T.; Fernandez, C. Zinc oxide nanoparticles modified-carbon paste electrode used for the electrochemical determination of Gallic acid. *J Phys Conf Ser.* **2019**, *1310*. doi:10.1088/1742-6596/1310/1/012008.
64. Wardak, C.; Paczosa-Bator, B.; Malinowski, S. Application of cold plasma corona discharge in preparation of laccase-based biosensors for dopamine determination. *Materials Science and Engineering: C.* **2020**, *116*, 1–10. doi:10.1016/j.msec.2020.111119.
65. Azad, M.S. Experiences of Domestic Abuse within the South Asian Community. *J glob faultlines.* **2021**, *8*, 50–68.
66. Gil, D.M.A.; Rebelo, M.J.F. Evaluating the antioxidant capacity of wines: A laccase-based biosensor approach. *Eur Food Res Technol.* **2010**, *231*, 303–308.
67. Gahlaut, A.; Gothwal, A.; Chhillar, K.; Hooda, V. Electrochemical Biosensors for Determination of Organophosphorus Compounds : Review. *Sci Res J.* **2012**, *1*, 1-8.
68. de Oliveira Neto, J.R.; Rezende, S.G.; Lobón, G.S.; Garcia, T.A.; Macedo, I.Y.L.; Garcia, L.F.; Alves, V.F.; Torres, I.M.S.; Santiago, M.F.; Schmidt, F.; de Souza Gil, E. Electroanalysis and laccase-based biosensor on the determination of phenolic content and antioxidant power of honey samples. *Food Chem.* **2017**, *237*, 1118–1123. doi:10.1016/j.foodchem.2017.06.010.
69. Tomac, I., Šeruga, M., Labuda, J. Evaluation of antioxidant activity of chlorogenic acids and coffee extracts by an electrochemical DNA-based biosensor. *Food Chem.* **2020**, *325*, 126787. doi:10.1016/j.foodchem.2020.126787.
70. Sharma, S.; Byrne, H.; O'Kennedy, R.J. Antibodies and antibody-derived analytical biosensors. *Essays Biochem.* **2016**, *60*, 9-18. doi: 10.1042/EBC20150002
71. Ye, Y.; Sun, X.; Zhang, Y.; Han, X.; Sun, X. A novel cell-based electrochemical biosensor based on MnO₂ catalysis for antioxidant activity evaluation of anthocyanins. *Biosens Bioelectron.* **2022**, *202*, 113990. doi:10.1016/j.bios.2022.113990
72. Narita, F.; Wang, Z.; Kurita, H.; Li, Z.; Shi, Y.; Jia, Y.; Soutis, C. A Review of Piezoelectric and Magnetostrictive Biosensor Materials for Detection of COVID-19 and Other Viruses. *Adv. Mater.* **2021**, *33*, 2005448.

73. Damborský, P.; Švitel, J.; Katrlík, J. Optical Biosensors. *Essays Biochem.* **2016**, *60*, 91–100. doi:10.1042/EBC20150010

74. Hackney, A.C. Exercise, Sport, and Bioanalytical Chemistry; Elsevier: Amsterdam, The Netherlands, 2016.

75. Dzyadevych, S.; Jaffrezic-Renault, N. Conductometric Biosensors. In Biological Identification; Elsevier: Amsterdam, The Netherlands, 2014; pp. 153–193.

76. Fu, Y.; You, Z.; Xiao, A.; Liu, L.; Zhou, W. Electrochemical evaluation of the antioxidant capacity of natural compounds on glassy carbon electrode modified with guanine-, polythionine-, and nitrogen-doped graphene. *Open Chem.* **2020**, *18*, 1054–1063. doi:10.1515/chem-2020-0157.

77. Cesewski, E.; Johnson, B.N. Electrochemical biosensors for pathogen detection. *Biosens Bioelectron.* **2020**, *159*, 112214. doi:10.1016/j.bios.2020.112214.

78. Jin, X.; Liu, C.; Xu, T.; Su, L.; Zhang, X. Artificial intelligence biosensors: Challenges and prospects. *Biosens Bioelectron.* **2020**, *165*, 112412. doi:10.1016/j.bios.2020.112412.

79. Cetó, X.; Céspedes, F.; Pividori, M.I.; Gutiérrez, J.M.; Del Valle, M. Resolution of phenolic antioxidant mixtures employing a voltammetric bio-electronic tongue. *Analyst.* **2012**, *137*, 349–356. doi:10.1039/c1an15456g.

80. Vlomidis, Y.; Gualandi, I.; Tonelli, D. Amperometric biosensors based on reduced GO and MWCNTs composite for polyphenols detection in fruit juices. *J Electroanal Chem.* **2017**, *799*, 285–292.

81. Nadifiyine, S.; Calas-Blanchard, C.; Amine, A.; Marty, J.L. Tyrosinase Biosensor Used for the Determination of Catechin Derivatives in Tea: Correlation with HPLC/DAD Method. *Food Nutr Sci.* **2013**, *4*, 108–118. doi:10.4236/fns.2013.41015.

82. Sýs, M.; Pekec, B.; Kalcher, K.; Vytřas, K. Amperometric enzyme carbon paste-based biosensor for quantification of hydroquinone and polyphenolic antioxidant capacity. *Int J Electrochem Sci.* **2013**, *8*, 9030–9040.

83. Munteanu, I.G.; Apetrei, C. Tyrosinase-Based Biosensor—A New Tool for Chlorogenic Acid Detection in Nutraceutical Formulations. *Materials (Basel).* **2022a**, *15*, 3221. doi:10.3390/ma15093221.

84. Bounegru, A.V.; Apetrei, C. Laccase and tyrosinase biosensors used in the determination of hydroxycinnamic acids. *Int. J. Mol. Sci.* **2021**, *22*, 4881. doi:10.3390/ijms22094811.

85. Kadam, A.A.; Saratale, G.D.; Ghodake, G.S.; Saratale, R.G.; Shahzad, A.; Magotra, V.K.; Kumar, M.; Palem, R.R.; Sung, J.S. Recent Advances in the Development of Laccase-Based Biosensors via Nano-Immobilization Techniques. *Chemosensors.* **2022**, *10*, 1–27. doi:10.3390/chemosensors10020058.

86. Rodríguez-Delgado, M.M.; Alemán-Nava, G.S.; Rodríguez-Delgado, J.M.; Dieck-Assad, G.; Martínez-Chapa, S.O.; Barceló, D.; Parra, R. Laccase-based biosensors for detection of phenolic compounds. *TrAC - Trends Anal Chem.* **2015**, *74*, 21–45. doi:10.1016/j.trac.2015.05.008.

87. Munteanu, I.G.; Apetrei, C. Assessment of the Antioxidant Activity of Catechin in Nutraceuticals: Comparison between a Newly Developed Electrochemical Method and Spectrophotometric Methods. *Int J Mol Sci.* **2022b**, *23*, 8110. doi:10.3390/ijms23158110.

88. Yang, Y.; Zeng, H.; Zhang, Q.; Bai, X.; Liu, C.; Zhang, Y.H. Direct electron transfer and sensing performance for catechin of nano-gold particles-polymer nano-composite with immobilized Laccase. *Chem Phys Lett.* **2016**, *658*, 259–269. doi:10.1016/j.cplett.2016.06.063.

89. Steevensz, A.; Villegas L.G.C.; Feng, W.; Taylor, K.E.; Bewtra, J.K.; Biswas, N. Soybean peroxidase for industrial wastewater treatment: a mini review. *J Environ Eng Sci.* **2014**, *9*, 181–186. doi:10.1680/jees.13.00013.

90. Neumann, B.; Wollenberger, U. Electrochemical Biosensors Employing Natural and Artificial Heme Peroxidases on Semiconductors. *Sensors* **2020**, *20*, 3692. <https://doi.org/10.3390/s20133692>

91. Smutok, O.; Kavetskyy, T.; Prokopiv, T.; Serkiz, R.; Šauša, O.; Novák, I.; Švajdlenková, H.; Maťko, I.; Gonchar, M.; Katz, E. Biosensor Based on Peroxidase-Mimetic Nanozyme and Lactate Oxidase for Accurate L-Lactate Analysis in Beverages. *Biosensors.* **2022**, *12*, 1042. doi:10.3390/bios12111042.

92. Shin, J.H.; Lee, M.J.; Choi, J.H.; Song, J.; Kim, T.H.; Oh, B.K. Electrochemical H₂O₂ biosensor based on horseradish peroxidase encapsulated protein nanoparticles with reduced graphene oxide-modified gold electrode. *Nano converg.* **2020**, *7*, 1–8, <https://doi.org/10.1186/s40580-020-00249-0>.

93. Wu, L.; Yin, W.; Tang, K.; Li, D.; Shao, K.; Zuo, Y.; Ma, J.; Liu, J.; Han, H. Enzymatic biosensor of horseradish peroxidase immobilized on Au-Pt nanotube/Au-graphene for the simultaneous determination of antioxidants. *Anal Chim Acta.* **2016**, *933*, 89–96. doi:10.1016/j.aca.2016.06.020.

94. Liu, J. DNA-stabilized, fluorescent, metal nanoclusters for biosensor development. *TrAC - Trends Anal Chem.* **2014**, *58*, 99–111. doi:10.1016/j.trac.2013.12.014.

95. Gupta, N.; Renugopalakrishnan, V.; Liepmann, D.; Paulmurugan, R.; Malhotra, B.D. Cell-based biosensors: Recent trends, challenges and future perspectives. *Biosens Bioelectron.* **2019**, *141*, 111435. doi:10.1016/j.bios.2019.111435.

96. Inda, M.E.; Mimee, M.; Lu, T.K. Cell-based biosensors for immunology, inflammation, and allergy. *J Allergy Clin Immunol.* **2019**, *144*, 645–647. doi: 10.1016/j.jaci.2019.07.024.

97. Ge, Q.; Ge, P.; Jiang, D.; Du, N.; Chen, J.; Yuan, L.; Yu, H.; Xu, X.; Wu, M.; Zhang, W.; Zhou, G. A novel and simple cell-based electrochemical biosensor for evaluating the antioxidant capacity of *Lactobacillus plantarum* strains isolated from Chinese dry-cured ham. *Biosens Bioelectron.* **2018**, *99*, 555–563. doi:10.1016/j.bios.2017.08.037.

98. Lindeman, N.I.; Cagle, P.T.; Aisner, D.L.; Arcila, M.E.; Beasley, M.B.; Bernicker, E.H.; Colasacco, C.; Dacic, S.; Hirsch, F.R.; Kerr K, et al. Updated molecular testing guideline for the selection of lung cancer patients for treatment with targeted tyrosine kinase inhibitors guideline from the college of American pathologists, the international association for the study of lung cancer, and the a. *Arch Pathol Lab Med.* **2018**, *142*, 321–346. doi:10.5858/arpa.2017-0388-CP.

99. Iswantini, D.; Weniarti.; Nurhidayat, N.; Abidin, Z.; Trividila. Antioxidant biosensor based on superoxide dismutase from Indonesian microbes immobilized in Indonesian natural zeolite. *J Appl Pharm Sci.* **2019**, *9*, 104–109. doi:10.7324/JAPS.2019.90413.

100. Iswantini, D.; Nurhidayat, N.; Trividila; Widiyatmoko, O. Activity and stability of uricase from *Lactobacillus plantarum* immobilized on natural zeolite for uric acid biosensor. *Pak J Biol Sci.* **2014**, *17*, :277–281. doi:10.3923/pjbs.2014.277.281

101. Febryanti, A.; Mulijani, S.; Iswantini D. Antioxidant biosensor based on *Deinococcus radiodurans* biofilm immobilized on screen printed carbon electrode (SPCE) surface. *IJRASB*, **2018**, *5*, 1–7.

102. Reder-Christ, K.; Bendas, G. Biosensor applications in the field of antibiotic research-a review of recent developments. *Sensors.* **2011**, *11*, 9450–9466. doi:10.3390/s111009450.

103. Liang, G.; Man, Y.; Li, A.; Jin, X.; Liu, X.; Pan, L. DNAzyme-based biosensor for detection of lead ion: A review. *Microchem J.* **2017**, *131*, 145–153. doi:10.1016/j.microc.2016.12.010.

104. Saidur, M.R.; Aziz, A.R.A.; Basirun, W.J. Recent advances in DNA-based electrochemical biosensors for heavy metal ion detection: A review. *Biosens Bioelectron.* **2017**, *90*, 125–139. doi:10.1016/j.bios.2016.11.039.

105. Dehghani, S.; Nosrati, R.; Yousefi, M.; Nezami, A.; Soltani, F.; Taghdisi, S.M.; Abnous, K.; Alibolandi, M.; Ramezani, M. Aptamer-based biosensors and nanosensors for the detection of vascular endothelial growth factor (VEGF): A review. *Biosens Bioelectron.* **2018**, *110*, 23–37. doi:10.1016/j.bios.2018.03.037.

106. Wu, Y.; Gao, H.; Wang, Y.; Peng, Z.; Guo, Z.; Ma, Y.; Zhang, R.; Zhang, M.; Wu, Q.; Xiao, J.; Zhong, Q. Effects of different extraction methods on contents, profiles, and antioxidant abilities of free and bound phenolics of *Sargassum polycystum* from the South China Sea. *J Food Sci.* **2022**, *87*, 968–981. doi:10.1111/1750-3841.16051.

107. Thapa, K.; Liu, W.; Wang, R. Nucleic acid-based electrochemical biosensor: Recent advances in probe immobilization and signal amplification strategies. *Wiley Interdiscip Rev Nanomedicine Nanobiotechnology.* **2022**, *14*, 1–25. doi:10.1002/wnan.1765.

108. Mehlhorn, A.; Rahimi, P.; Joseph, Y. Aptamer-based biosensors for antibiotic detection: A review. *Biosensors.* **2018**, *8*, 54. doi:10.3390/bios8020054.

109. Li, N.; Du, M.; Liu, Y.; Ji, X.; He, Z. Multipedal DNA Walker Biosensors Based on Catalyzed Hairpin Assembly and Isothermal Strand-Displacement Polymerase Reaction for the Chemiluminescent Detection of Proteins. *ACS Sensors.* **2018**, *3*, 1283–1290. doi:10.1021/acssensors.8b00129.

110. Minagawa, H.; Kataoka, Y.; Kuwahara, M.; Horii, K.; Shiratori, I.; Waga, I. A high affinity modified DNA aptamer containing base-appended bases for human β -defensin. *Anal Biochem.* **2020**, *594*, 8–11. doi:10.1016/j.ab.2020.113627.

111. Yang, H.; Zhou, Y.; Liu, J. G-quadruplex DNA for construction of biosensors. *TrAC - Trends Anal Chem.* **2020**, *132*, 116060. doi:10.1016/j.trac.2020.116060.

112. Xue, C.; Zhang, S.; Yu, X.; Hu, S.; Lu, Y.; Wu, Z.S. Periodically Ordered, Nuclease-Resistant DNA Nanowires Decorated with Cell-Specific Aptamers as Selective Theranostic Agents. *Angew Chemie - Int Ed.* **2020**, *59*, 17540–17547. doi:10.1002/anie.202004805.

113. Liu, J. DNA-stabilized, fluorescent, metal nanoclusters for biosensor development. *TrAC - Trends Anal Chem.* **2014**, *58*, 99–111. doi:10.1016/j.trac.2013.12.014
114. Diculescu, V.C.; Chiorcea-Paquim, A.M.; Oliveira-Brett, A.M. Applications of a DNA-electrochemical biosensor. *TrAC - Trends Anal Chem.* **2016**, *79*, 23–36. doi:10.1016/j.trac.2016.01.019.
115. Barroso, M.F.; Delerue-Matos, C.; Oliveira, M.B.P.P. Electrochemical evaluation of total antioxidant capacity of beverages using a purine-biosensor. *Food Chem.* **2012**, *132*, 1055–1062. doi:10.1016/j.foodchem.2011.10.072.
116. Peng, H.; Wang, D.; Li, M.; Zhang, L.; Liu, M.; Fu, S. N-P-Zn-containing 2D supermolecular networks grown on MoS₂ nanosheets for mechanical and flame-retardant reinforcements of polyacrylonitrile fiber. *Chem Eng J.* **2019**, *372*, 873–885. doi:10.1016/j.cej.2019.04.209.
117. Tang, P.; Tang, X.; Mei, S.; Xie, Y.; Liu, L.; Ren, L. Electrochemical antioxidant screening and evaluation based on guanine and chitosan immobilized MoS₂ nanosheet modified glassy carbon electrode (guanine/CS/MoS₂/GCE). *Open Chem.* **2020**, *18*, 1–9. doi:10.1515/chem-2020-0003.
118. Tomac, I.; Šeruga, M.; Labuda, J. Evaluation of antioxidant activity of chlorogenic acids and coffee extracts by an electrochemical DNA-based biosensor. *Food Chem.* **2020**, *325*, 126787. doi:10.1016/j.foodchem.2020.126787.
119. Ensafi, A.A.; Kazemnadi, N.; Amini, M.; Rezaei, B. Impedimetric DNA-biosensor for the study of dopamine induces DNA damage and investigation of inhibitory and repair effects of some antioxidants. *Bioelectrochemistry.* **2015**, *104*, 71–78. doi:10.1016/j.bioelechem.2015.03.008
120. Chiorcea-Paquim, A.M.; Enache, T.A.; De Souza Gil, E.; Oliveira-Brett, A.M. Natural phenolic antioxidants electrochemistry: Towards a new food science methodology. *Compr Rev Food Sci Food Saf.* **2020**, *19*, 1680–1726. doi:10.1111/1541-4337.12566.
121. Manoranjitham, J.J.; Narayanan, S.S. Electrochemical sensor for determination of butylated hydroxyanisole (BHA) in food products using poly O-cresolphthalein complexone coated multiwalled carbon nanotubes electrode. *Food Chem.* **2021**, *342*, 128246. doi:10.1016/j.foodchem.2020.128246.
122. Fan, L.; Kan, X. Sensitive detection of butylated hydroxyanisole based on free-standing paper decorated with gold and NiO nanoparticles. *Microchem J.* **2020**, *159*, 105511. doi:10.1016/j.microc.2020.105511.
123. Ziyatdinova, G.; Guss, E.; Budnikov, H. Amperometric sensor based on MWNT and electropolymerized carminic acid for the simultaneous quantification of TBHQ and BHA. *J Electroanal Chem.* **2020**, *859*, 113885. doi:10.1016/j.jelechem.2020.113885.
124. Motia, S.; Bouchikhi, B.; El Bari, N. An electrochemical molecularly imprinted sensor based on chitosan capped with gold nanoparticles and its application for highly sensitive butylated hydroxyanisole analysis in foodstuff products. *Talanta.* **2020**, *223*, 121689. doi:10.1016/j.talanta.2020.121689.
125. Balram, D.; Lian, K.Y.; Sebastian, N.; Rasana, N. Ultrasensitive detection of cytotoxic food preservative tert-butylhydroquinone using 3D cupric oxide nanoflowers embedded functionalized carbon nanotubes. *J Hazard Mater.* **2021**, *406*, 124792. doi:10.1016/j.jhazmat.2020.124792
126. Yan, F.; Wang, M.; Jin, Q.; Zhou, H.; Xie, L.; Tang, H.; Liu, J. Vertically-ordered mesoporous silica films on graphene for anti-fouling electrochemical detection of tert-butylhydroquinone in cosmetics and edible oils. *J Electroanal Chem.* **2021**, *881*, 114969. doi:10.1016/j.jelechem.2020.114969.
127. Gao, Q.; Zang, Y.; Zhang, Y.; Xie, J.; Li, J.; Gao, J.; Xue, H. Composite polymerized molecular imprinting membrane-based electrochemical sensor for sensitive determination of curcumin by using 4-pentenoyl-aminoacyl-chitosan oligosaccharide as functional monomer oligomer. *J Electroanal Chem.* **2020**, *879*, 114793. doi:10.1016/j.jelechem.2020.114793.
128. Vinothkumar, V.; Sangili, A.; Chen, S.M.; Veerakumar, P.; Lin, K.C. Sr-Doped NiO₃ nanorods synthesized by a simple sonochemical method as excellent materials for voltammetric determination of quercetin. *New J Chem.* **2020**, *44*, 2821–2832. doi:10.1039/c9nj05660b.
129. Alipour, S.; Azar, P.A.; Husain, S.W.; Rajabi, H.R. Determination of Rosmarinic acid in plant extracts using a modified sensor based on magnetic imprinted polymeric nanostructures. *Sensors Actuators, B Chem.* **2020**, *323*, 128668. doi:10.1016/j.snb.2020.128668.
130. Lu, Y.; Hu, J.; Zeng, Y.; Zhu, Y.; Wang, H.; Lei, X.; Huang, S.; Guo, L.; Li, L. Electrochemical determination of rutin based on molecularly imprinted poly (ionic liquid) with ionic liquid-graphene as a sensitive element. *Sensors Actuators, B Chem.* **2020**, *311*, 127911. doi:10.1016/j.snb.2020.127911.

131. El Jaouhari, A.; Yan, L.; Zhu, J.; Zhao, D.; Khan, M.Z.H.; Liu, X. Enhanced molecular imprinted electrochemical sensor based on zeolitic imidazolate framework/reduced graphene oxide for highly recognition of rutin. *Anal Chim Acta*. **2020**, *1106*, 103–114. doi:10.1016/j.aca.2020.01.039.
132. Thangavel, B.; Lingagounder, J.; Berchmans, S.; Ganesh, V. Electrochemical grafting of 4-aminobenzoic acid onto toray carbon – Interfacial investigation and fabrication of non-enzymatic bilirubin sensor. *Sensors Actuators, B Chem.* **2021**, *344*, 130292. doi:10.1016/j.snb.2021.130292
133. Bettazzi, F.; Ingrosso, C.; Sfragano, P.S.; Pifferi, V.; Falciola, L.; Curri, M.L.; Palchetti, I. Gold nanoparticles modified graphene platforms for highly sensitive electrochemical detection of vitamin C in infant food and formulae. *Food Chem.* **2021**, *344*, 128692. doi:10.1016/j.foodchem.2020.128692.
134. Xiao, X.; Zhang, Z.; Nan, F.; Zhao, Y.; Wang, P.; He, F.; Wang, Y. Mesoporous CuCo₂O₄ rods modified glassy carbon electrode as a novel non-enzymatic amperometric electrochemical sensors with high-sensitive ascorbic acid recognition. *J Alloys Compd.* **2021**, *852*, 157045. doi:10.1016/j.jallcom.2020.157045.
135. Matsunaga, T.; Kondo, T.; Shitanda, I.; Hoshi, Y.; Itagaki, M.; Tojo, T.; Yuasa, M. Sensitive electrochemical detection of L-Cysteine at a screen-printed diamond electrode. *Carbon N Y.* **2021**, *173*, 395–402. doi:10.1016/j.carbon.2020.10.096.
136. Maraldi, V.A.; Colmenares, Y.N.; Barbosa, P.P.F., Mastelaro, V.; do Carmo, D.R. Graphene Oxide as a Platform for Copper Pentacyanonitrosylferrate Nanoparticles and their Behavior in the Electro-oxidation of N-Acetylcysteine. *Electroanalysis*. **2020**, *32*, 1408–1416. doi:10.1002/elan.201900493.

Disclaimer/Publisher's Note: The statements, opinions and data contained in all publications are solely those of the individual author(s) and contributor(s) and not of MDPI and/or the editor(s). MDPI and/or the editor(s) disclaim responsibility for any injury to people or property resulting from any ideas, methods, instructions or products referred to in the content.