

## Article

# Photothermal off-resonance tapping for rapid and gentle atomic force imaging of live cells

Adrian P. Nievergelt<sup>1</sup> , Charlene Brillard<sup>1</sup>, Haig A. Eskandarian<sup>1,2</sup> , John D. McKinney<sup>2</sup>  and Georg E. Fantner<sup>1\*</sup> 

<sup>1</sup> Laboratory for Bio- and Nano-Instrumentation, Swiss Federal Institute of Technology Lausanne (EPFL), Lausanne 1015, Switzerland. georg.fantner@epfl.ch

<sup>2</sup> UPKIN, Swiss Federal Institute of Technology Lausanne (EPFL), Lausanne 1015, Switzerland.

\* Correspondence: georg.fantner@epfl.ch; Tel.: +41 21 69 36431

**Abstract:** Imaging living cells by atomic force microscopy (AFM) promises not only high-resolution topographical data, but additionally, mechanical contrast, which are not obtainable with other microscopy techniques. Such imaging is however challenging, as cells need to be measured with low interaction forces to prevent either deformation or detachment from the surface. Off-resonance modes which periodically probe the surface have been shown to be advantageous, as they provide excellent force control combined with large amplitudes, which help reduce lateral force interactions. However, the low actuation frequency in traditional off-resonance techniques limits the imaging speed significantly. Using photothermal actuation, we probe the surface by directly actuating the cantilever. Due to the much smaller mass that needs to be actuated, the achievable measurement frequency is increased by two orders of magnitude. Additionally, photothermal off-resonance tapping retains the precise force control of conventional off-resonance modes and is therefore well suited to gentle imaging. Here we show how photothermal off-resonance tapping can be used to study live cells by AFM. As an example of imaging mammalian cells, the initial attachment, as well as long term detachment of a human thrombocytes are presented. The membrane disrupting effect of the antimicrobial peptide CM-15 is shown on the cell wall of *E. coli*. Finally, the dissolution of the cell wall of *B. subtilis* by lysozyme is shown. Taken together, these evolutionarily disparate forms of life exemplify the usefulness of PORT for live cell imaging in a multitude of biological disciplines.


**Keywords:** High-speed atomic force microscopy, live cell imaging, antimicrobial peptide, thrombocytes, bacterial imaging, cell lysis

## 1. Introduction

The mechanical nature of atomic force microscopy (AFM) makes it a powerful complementary technique to optical imaging for live-cell experiments, because it can offer nanometer resolution [1,2] as well as mechanical information of the sample [3–5]. While there are a wide range of imaging modalities available for AFM, only a subset is usable for live cell imaging. The main difficulties in observing living cells is that they are either very soft or not well attached to the underlying substrate [6,7]. Classical AFM modes like amplitude modulation or contact mode tend to require the sample to be well attached when working in liquid due to large interaction forces, both normal [8,9] as well as lateral [10]. Additionally, neither of these modes on their own can provide reliable mechanical information on the sample [11]. Often, force-distance based modes provide both better force control as well as less lateral interaction during scanning, although usually at the cost of imaging speed when compared to resonant modes [12].

Force-volume, where individual force ramps are done in multiple points of the sample [13,14], provides the best force control as well as reliable mechanical property measurements. However, the imaging rate in the order of tens of minutes to hours per image is by far too slow to observe dynamic effects on cells. Off-resonant modes (Pulsed force mode, PeakForce, Jumping mode, HybriD, QI Mode)

on the other hand modulate the tip-sample distance periodically at a frequency much lower than the resonance frequency of the cantilever to obtain force interactions at the modulation rate [15,16]. Previously, off-resonant modes have been applied with great success to a wide range of problems in molecular and cell biology, such as molecular recognition [17–19], cell mechanics [20,21], host-pathogen interaction [22]. The maximum force during such an interaction is then used as the main feedback variable during scanning. In addition to increasing the image acquisition speed, these modes are well suited to obtain mechanical contrast since the resulting periodic force interactions can be analysed in real time. Furthermore, off-resonance modes also tend to be easier to use, especially in liquid, since they can be made immune to set-point drift. In most resonant modes, the free amplitude becomes unobservable once in feedback. Since the amplitude is being actively controlled for, the observed amplitude will always correspond to the set-point, as long as that set-point amplitude is smaller than the free amplitude and the piezo is in range. However, the free amplitude can change almost arbitrarily. The free amplitude can therefore only be checked by periodically withdrawing from the surface. In contrast, the interaction in off-resonant modes can be fully observed in every cycle, since the cantilever is disengaged from contact for most of the measurement cycle, allowing for the establishment of a per-cycle tip-sample force free baseline value. These properties make off-resonance modes an excellent mode to study live cells [2,23].



**Figure 1.** Working principle and implementation of photothermal off-resonance tapping (PORT). (a) Conventionally, off-resonance modes use the axial motion of the scanner to periodically probe the surface in a controlled fashion [24]. The resonances of those scanners limit the probing frequency to about 2 kHz for tube scanners and to about 12 kHz for a high-rate (HR) scanner. (b) Schematic of PORT setup and operation: the cantilever is driven with a secondary driving laser while a dedicated PORT controller is used to extract the tip-sample interaction from the cantilever deflection signal. (c) The heating induced by the drive laser (thermal distribution along the cantilever indicated top right) causes a differential expansion in the cantilever which leads to controlled bending. Smaller cantilevers are more suited to PORT, both due to increased driving efficiency as well as higher resonance frequency. Examples of live cells scanned in PORT are (d) human thrombocytes, (e) gram-negative bacteria (*E. coli*) as well as (f) gram-positive bacteria (*B. subtilis*). All scale bars are 2  $\mu\text{m}$ .


However, even at the improved imaging speeds, it often takes on the order of several minutes to acquire an image, especially on difficult samples like live cells [25].

The speed limit in classical off-resonance modes stems from the fact that the tip-sample modulation frequency has to be kept significantly below the first z-resonance of the scanner in order to produce a controlled motion (see figure 1a). To bypass the inertial effect which causes the scanner resonances, we actuate the cantilever directly using a laser beam (figure 1b). By this method, generally known as photothermal excitation [26–30], the mass which needs to be actuated is reduced to only the cantilever and the tip (figure 1c). It is therefore possible to increase the rate at which the surface is interrogated by over two orders of magnitude, while maintaining the ability to use large amplitudes and image with a controlled force. Previously, we have shown how this technique, called photothermal off-resonance tapping (PORT), can be used to measure the self-assembly of proteins in real time [31]. Here we show how the same technique enables robust imaging of living cells at substantially higher speeds than traditionally possible (figure 1d-f). One of the key strengths of PORT is the ability to use large oscillation amplitudes of easily more than 50 nm, even when operating in liquid, while at the same time maintaining a small force interaction. Such large amplitudes reduce the probability that a large change in sample height will quench the whole oscillation. In that occurrence, very large lateral force interactions are expected while the tip stays in contact with the sample for a prolonged amount of time. Such large changes in height are common with big samples such as living cells. As a consequence, both bacteria and eukaryotes alike can be maintained in physiological conditions and rapid events measured. Due to the slow time scales involved, traditional time-lapse AFM techniques used to study cells [25] often do not resolve such processes properly.

## 2. Results


### 2.1. Thrombocyte imaging

Thrombocytes are a component in blood which are involved in the formation of clots that stop bleeding from wounds. They are nucleus-free cells produced in bone marrow [32,33]. Due to their relatively small size in the order of 3–10  $\mu\text{m}$ , they are a challenging sample for optical microscopy to resolve spatially [34]. Due to the large oscillation amplitudes in PORT and the subsequently low lateral forces, the initial adhesion and spreading of a thrombocyte can be observed (see figure 2). In just a few seconds, the thrombocyte, which still retains most of its lentil shape, forms pseudopodia which act as additional anchors for the cell while it spreads on the glass surface.



**Figure 2.** The attachment of a human thrombocyte onto a glass cover slip measured at 40 kHz photothermal off-resonance tapping AFM. Pseudopodia can be seen forming in (a), which keep expanding (c)–(f) as the thrombocyte. The shadow visible to the left of the cell in (a)–(c) is due to insufficient tip length and subsequent force interaction with the cantilever. Scale bars are 2  $\mu\text{m}$  and height of cell in (a) is 1.6  $\mu\text{m}$ . Imaging rate is 16 s/frame, 4 lines/s.

In addition to short term cell dynamics, like attachment, we have used PORT for time lapse imaging on thrombocyte cells that are fully spread on a glass surface and subsequently detaches (see figure 3, supplementary movie 1).




**Figure 3.** Time-lapse measurement of thrombocyte detachment from glass surface. While initially spread on the surface, the cell over time contracts (c) and changes into a more spherical shape, then eventually detaches from the surface in (k). Imaging rate is 23 s per frame, 11 lines/s.

Initially, the thrombocytes are mostly static (figure 3a,b). Over time the cell starts contracting and gains in height appreciatively (figure 3c-d). Over the next minute the cell continuously reduces the area which is in contact with the glass surface and returns to a more lentil shaped form. During this process the membrane at the edges of the cells retracts, but patches of cellular matter are seen to stay on the glass. Eventually the attachment of the cell to the glass surface is insufficient to hold the cell down. Subsequently the thrombocyte is swiped away by the AFM tip. A thin layer of presumably membrane material remains where the cell was originally attached on the surface. The layer is spotty towards the borders of where the cell used to reside and continuous in the center.

## 2.2. Bacterial imaging

In order to show the suitability of PORT to imaging processes on bacterial cells, we imaged membrane disruption on both a gram-negative bacterium (*E. coli*) as well as a gram-positive bacterium (*B. subtilis*).

We have previously shown that the antimicrobial peptide CM-15 causes surface roughening on *E. coli* cells [35,36]. Here we repeat this experiment using PORT. In contrast to our previous work, here we use a minimal growth medium, which prolongs cell viability significantly as compared to suspending them in deionized water. However, due to the ionic strength of the growth medium, the immobilization with poly-L-lysine is significantly weaker and scanning the cells in amplitude modulation often causes them to be detached from the surface [37]. Using PORT, we are able to scan live *E. coli* in unsupplemented as well as supplemented growth medium over extended periods of time.




**Figure 4.** Membrane disruption on *E. coli* due to attack by the antimicrobial peptide CM15 (added at  $t = 0$  s). The cell membrane is significantly deformed over time. Notably, the central constriction where the cell has started dividing becomes more pronounced. (a) Height image. (b) Force error image. Images are taken at 12 s/frame, 10.3 lines/s, measured at 16kHz PORT rate.

Figure 4 shows a time-lapse sequence in which CM-15 modifies the cell wall (see supplementary movie 2). Over time the cell seems to shrivel up while initially maintaining turgor pressure[38]. The effect is especially pronounced at the constriction where the cell has started dividing. A more pronounced buckling can be observed at the location of this constriction, which might result from the stress geometry of the constriction itself.

Compared to gram-negative bacteria, the cell wall of gram-positive bacteria like *B. subtilis* is significantly thicker, estimated to be about 30 nm to 60 nm thick, a large fraction of which is composed of a peptidoglycan layer that forms the outer cell wall [39]. This layer forms the load bearing support for the bacterial cell. We perturbed the peptidoglycan layer by treating cells with lysozyme, an enzyme that hydrolyzes cross-linked peptidoglycan. Lysozomal activity in eukaryotic internalized vesicles is an innate part of mammalian immune reaction. Lysozyme hydrolyses peptidoglycan, compromising bacterial structural integrity, causing morphological bulging [40] and ultimately causing bacterial lysis.

We have acquired time lapse AFM images of the lysis process on the cell wall of *B. subtilis* bacteria (see figure 5, supplementary movie 3). After 10 min of imaging, we added a substantial amount of lysozyme (20  $\mu$ L of 500  $\mu$ g  $\text{mL}^{-1}$  lysozyme in 50 mM TRIS) to the liquid cell of the AFM.



**Figure 5.** The effect of lysozyme on the cell wall of *B. subtilis* in growth medium. Lysozyme is added at  $t = 10$  min. a) Height image. b) The error image shows how initial scars in the cell wall (circled marks) of the bacterium, which then expand to patches (square marks) where the peptidoglycan layer is partially removed, followed by rapid loss of turgor pressure and subsequent cell death. The PORT adhesion increases slightly as more of the peptidoglycan layer is hydrolyzed. Notable are dark patches visible in the height channel (square marks), which have lower adhesion where presumably most of the peptidoglycan has been stripped. Scan size  $3 \mu\text{m}$ , 7 lines/s, measured at 25kHz PORT rate.

A few seconds after injection ( $t = 10$  min), we see a slight roughening of the cell wall, together with a widening of already present scars on the outside of the bacterium (see figure 5b, circled marks). An initial small decrease in cell diameter is also apparent, suggesting a loss of turgor pressure. The scars expand over the next few minutes, until whole patches of peptidoglycan are dissolved and the underlying layer is exposed (see figure 5a square marks). With ongoing dissolution of the peptidoglycan layer, the cell loses its structural integrity and deflates over the next minutes until only a very soft remainder of the sacculus is present which is readily displaced by the tip.

### 3. Discussion

Here, we have shown the advantages of using PORT for scanning live bacterial and eucariotic cells. We have demonstrated intrinsic cell behaviours with the attachment and detachment of human thrombocyte cells as well as reactions of cells due to externally induced factors.

The main benefit of PORT for live-cell imaging is the ability to combine the high speed of resonant tapping with the large oscillation amplitudes that allow for a big periodic clearance from the sample which in turn reduces lateral forces that often displace high, weakly bound samples. Additionally, mechanical information can be extracted from the measurement at no added complexity to the user. Furthermore, the direct observability of the force interaction and the de-facto immunity to changes in the tip and the cantilever—both of which are major problems in resonant tapping—making fully automated measurements possible, which enables investigations of problems that are not addressable with conventional amplitude modulation AFM.

As with any atomic force microscopy imaging, cantilever selection is crucial to achieving good results. Cantilever interactions due to insufficient top length, such as are visible in figure 2a-c can lead to a wrong interpretation of height data and make volume calculations impractical. Focused electron

beam induced deposition of tips [41–43] can be used to grow micron-sized high aspect ratio tips on cantilevers that are supplied with short tips.

While PORT offers drastically improved feedback bandwidth compared to other ORT modes, it should be noted that achievable modulation amplitudes in liquid are usually limited to less than 150 nm. While this is more than sufficient for most samples, it might be insufficient to overcome sample adhesion forces in case of a very adherent substrate such as with polydimethylsiloxane (PDMS).

Even at the increased feedback speed, the large topography change which is especially pronounced at the border of a bacterial cell can cause the whole off-resonance amplitude to be quenched when scanning very fast. In this case the resulting large lateral forces can still result in the displacement of the cell. Standard techniques such as aligning the scan direction along the bacterium should still be employed. Finally, using pyramidal tips is recommended for imaging bacteria, as high aspect-ratio tips lead to sudden impacts with the cell.

The examples shown here are meant to demonstrate the versatility of the technique. Other examples of interesting dynamic processes include, but are not limited to thrombocyte migration, growth and division of morphologically different bacteria, pilus dynamics and mechanics, endospore formation or to study host-pathogen interactions. We believe that PORT opens up a large variety of possibilities in live cell imaging to study rapid behaviour of live cells and will continue the success of off-resonance AFM in cell biology.

## 4. Materials and Methods

### 4.1. Thrombocyte preparation

Thrombocyte samples were prepared by pipetting 10  $\mu$ L to 20  $\mu$ L of fresh blood, harvested from a lancette puncture on a finger tip, onto a clean 12 mm glass cover slip glued on a magnetic steel sample disk. The drop of blood is left to incubate for 5 min in order for thrombocytes to form initial attachment. The cover slips are then washed in Tyrode's Solution (3 mM HEPES, 137 mM NaCl, 4 mM  $\text{NaH}_2\text{PO}_4$ , 2.6 mM KCl, 1 mM  $\text{MgCl}_2$  and 1 mM Glucose) to remove any non-attached blood components. After washing, samples were immediately transferred to the AFM for imaging. Glass cover slips were cleaned by ultrasonication in 9:1, 1:1 and 1:9 mixtures of Chloroform:Ethanol, 1 min each.

### 4.2. Bacteria preparation

*E. coli* MG1665 were kindly donated by Yoshiko Miyahara from McKinney Lab at EPFL and *B. subtilis* 168 WT, donated by Stephan Gruber from UNIL. Cells were inoculated in LB Medium (*B. subtilis* and *E. coli* in figure 4) or in enriched M9 minimal medium (*E. coli* in figure 1e) in an orbital shaker at 37 °C and harvested in exponential phase. *E. coli* suspensions were subsequently washed three times in non-enriched M9 medium; *B. subtilis* was washed in M9 supplemented with 2 mM  $\text{MgSO}_4$  and 0.4 % glucose. Cleaned cultures were deposited on functionalized substrates. The functionalized surfaces were prepared by first ensuring a clean surface. Mica disks were cleaved with scotch tape, 12 mm glass cover slips were cleaned by 1 min  $\text{O}_2$ -plasma in a microwave plasma asher at 500 W (TePla 300). Both Mica and Glass surfaces were functionalized with poly-L-lysine (PLL) by immersing the substrate for half a minute in PLL solution (50  $\mu\text{g mL}^{-1}$  PLL in 10 mM TRIS, pH 8), then wicking off the drop with a paper tissue, rinsing thoroughly in deionized water and drying with a nitrogen stream.

Cells in about 30  $\mu$ L of solution were allowed to adhere to the surface between 5 min to 30 min and then rinsed in fresh buffer to remove floating cells. Prepared samples were immediately transferred to the AFM for imaging.

CM15, a peptide with the sequence KWKLFKKIGAVLKVL was bought from Genscript (Piscataway, NJ, USA) in lyophilized form and resuspended and diluted with millipore water at 1  $\text{mg mL}^{-1}$ . CM-15 was added to the experiment to achieve a final concentration of 50  $\mu\text{g mL}^{-1}$ .

Lysozyme was bought from Sigma-Aldrich (L6876) in lypophilized form, then resuspended in 50 mM TRIS at a concentration of  $5 \text{ mg mL}^{-1}$  and diluted to the final concentration of  $500 \mu\text{g mL}^{-1}$  before use.

#### 4.3. AFM Imaging

All imaging was performed on a home-built atomic force microscope, based on the Bruker MultiMode. A custom built drop-in replacement for the original microscope head, providing the means to use small cantilevers and photothermal excitation, was used and is described in detail elsewhere[29–31,44].

All imaging was done with either Olympus AC10DS (for thrombocytes) or Bruker FastScan-D cantilevers (for bacterial imaging). PORT amplitudes used were between 25 nm to 120 nm with typical setpoints between 0.5 nm to 5 nm, corresponding roughly to between 50 pN to 2000 pN.

PORT for platelet and *E. coli* imaging was implemented using a Nanoscope 5 controller with a modified PeakForce-HR workspace that allows splitting the modulation from the z-signal. A custom-built scaling and offset circuit was used to adapt output voltage levels as required by the head electronics. *B. subtilis* imaging was performed using a home-built AFM software based on a standalone FPGA (USB-7856R, National Instruments, Austin TX, USA) with software and hardware programming as described elsewhere [31] and using the amplifiers in a modified Nanoscope-IIIa controller (Digital Instruments, Santa Barbara CA, USA) that allow for input of external low-voltage scan signals. All images were aquired on a J-scanner (Bruker, Santa Barbara CA, USA) ( $120 \mu\text{m} \times 120 \mu\text{m} \times 5.2 \mu\text{m}$  range).

**Acknowledgments:** The authors would like to thank Shuiqing Hu and Chanmin Su from Bruker Nanotechnology for use of their controller and microscope for thrombocyte imaging. We would also like to thank Herman Schillers for help with thrombocyte sample preparation. This work was funded by the European Union's Seventh Framework Programme FP7/2007-2011 under grant agreement 286146, the European Union's Seventh Framework Programme FP7/2007-2013/ERC grant agreement 307338, and the Swiss National Science Foundation through grants 205321\_134786 and 205320\_152675, as well as by H2020 - EU Framework Programme for Research & Innovation (2014-2020); ERC-2017-CoG; InCell; Project number 773091. H.A.E. was supported by a European Molecular Biology Organization Long Term Fellowship (EMBO ALTF 191-2014 and aALTF 750-2016).

**Author Contributions:** A.P.N., C.B. and H.A.E and G.E.F conceived and designed the experiments; A.P.N and C.B performed the experiments; A.P.N. and C.B. analyzed the data; H.A.E. and J.D.M contributed reagents and materials; All authors contributed to writing the paper.

**Conflicts of Interest:** The authors declare no conflict of interest.

#### Abbreviations

The following abbreviations are used in this manuscript:

AFM	Atomic force microscopy
FPGA	Field Programmable Gate Array
HEPES	4-(2-hydroxyethyl)-1-piperazineethanesulfonic acid
PLL	Poly-L-lysine
PORT	Photothermal off-resonance tapping
TRIS	tris(hydroxymethyl)aminomethane

#### References

1. Henderson, E.; Haydon, P.G.; Sakaguchi, D.S. Actin filament dynamics in living glial cells imaged by atomic force microscopy. *Science (New York, N.Y.)* **1992**, *257*, 1944–6. doi:10.1126/SCIENCE.1411511.
2. Dufrène, Y.F.; Ando, T.; Garcia, R.; Alsteens, D.; Martinez-Martin, D.; Engel, A.; Gerber, C.; Müller, D.J. Imaging modes of atomic force microscopy for application in molecular and cell biology. *Nature Nanotechnology* **2017**, *12*, 295–307. doi:10.1038/nnano.2017.45.

3. Krottil, H.U.; Stifter, T.; Waschipky, H.; Weishaupt, K.; Hild, S.; Marti, O. Pulsed force mode: a new method for the investigation of surface properties. *Surface and Interface Analysis* **1999**, *27*, 336–340. doi:10.1002/(SICI)1096-9918(199905/06)27:5<336::AID-SIA512>3.0.CO;2-0.
4. Oberleithner, H.; Riethmuller, C.; Schillers, H.; MacGregor, G.A.; de Wardener, H.E.; Hausberg, M. Plasma sodium stiffens vascular endothelium and reduces nitric oxide release. *Proceedings of the National Academy of Sciences* **2007**, *104*, 16281–16286. doi:10.1073/pnas.0707791104.
5. Raman, A.; Trigueros, S.; Cartagena, A.; Stevenson, A.P.Z.; Susilo, M.; Nauman, E.; Contera, S.A. Mapping nanomechanical properties of live cells using multi-harmonic atomic force microscopy. *Nature Nanotechnology* **2011**, *6*, 809–814. doi:10.1038/nnano.2011.186.
6. Kailas, L.; Ratcliffe, E.; Hayhurst, E.; Walker, M.; Foster, S.; Hobbs, J. Immobilizing live bacteria for AFM imaging of cellular processes. *Ultramicroscopy* **2009**, *109*, 775–780. doi:10.1016/J.ULTRAMIC.2009.01.012.
7. Allison, D.P.; Sullivan, C.J.; Mortensen, N.P.; Retterer, S.T.; Doktycz, M. Bacterial immobilization for imaging by atomic force microscopy. *Journal of visualized experiments : JoVE* **2011**. doi:10.3791/2880.
8. Xu, X.; Carrasco, C.; de Pablo, P.J.; Gomez-Herrero, J.; Raman, A. Unmasking imaging forces on soft biological samples in liquids when using dynamic atomic force microscopy: a case study on viral capsids. *Biophysical journal* **2008**, *95*, 2520–8. doi:10.1529/biophysj.108.132829.
9. Guzman, H.V.; Perrino, A.P.; Garcia, R. Peak forces in high-resolution imaging of soft matter in liquid. *ACS Nano* **2013**, *7*, 3198–3204. doi:10.1021/nn4012835.
10. Putman, C.A.J.; Van Der Wetf, K.; De Grooth, B.G.; Hulst, N.F.V.; Greve, J. Tapping mode atomic force microscopy in liquid. *217*.
11. Nagao, E.; Dvorak, J.A. Phase Imaging by Atomic Force Microscopy: Analysis of Living Homoiothermic Vertebrate Cells. *Biophysical Journal* **1999**, *76*, 3289–3297. doi:10.1016/S0006-3495(99)77481-3.
12. Ortega-Esteban, A.; Horcas, I.; Hernando-Pérez, M.; Ares, P.; Pérez-Berná, A.J.; San Martín, C.; Carrascosa, J.L.; De Pablo, P.J.; Gómez-Herrero, J. Minimizing tip-sample forces in jumping mode atomic force microscopy in liquid. *Ultramicroscopy* **2012**, *114*, 56–61. doi:10.1016/j.ultramic.2012.01.007.
13. Solon, J.; Levental, I.; Sengupta, K.; Georges, P.C.; Janmey, P.A. Fibroblast Adaptation and Stiffness Matching to Soft Elastic Substrates. *Biophysical Journal* **2007**, *93*, 4453–4461. doi:10.1529/BIOPHYSJ.106.101386.
14. A-Hassan, E.; Heinz, W.F.; Antonik, M.D.; D'Costa, N.P.; Nageswaran, S.; Schoenenberger, C.A.; Hoh, J.H. Relative Microelastic Mapping of Living Cells by Atomic Force Microscopy. *Biophysical Journal* **1998**, *74*, 1564–1578. doi:10.1016/S0006-3495(98)77868-3.
15. Rosa-Zeiser, A.; Weilandt, E.; Hild, S.; Marti, O. The simultaneous measurement of elastic, electrostatic and adhesive properties by scanning force microscopy: pulsed-force mode operation. *Measurement Science and Technology* **1997**, *8*, 1333–1338. doi:10.1088/0957-0233/8/11/020.
16. De Pablo, P.J.; Colchero, J.; Gómez-Herrero, J.; Baró, A.M.; Gomez-Herrero, J.; Baro, a.M.; Gómez-Herrero, J.; Baró, A.M. Jumping mode scanning force microscopy. *Applied Physics Letters* **1998**, *73*, 3300–3302. doi:10.1063/1.122751.
17. Ludwig, M.; Dettmann, W.; Gaub, H. Atomic force microscope imaging contrast based on molecular recognition. *Biophysical Journal* **1997**, *72*, 445–448. doi:10.1016/S0006-3495(97)78685-5.
18. Alsteens, D.; Dupres, V.; Yunus, S.; Latgé, J.P.; Heinisch, J.J.; Dufrêne, Y.F. High-Resolution Imaging of Chemical and Biological Sites on Living Cells Using Peak Force Tapping Atomic Force Microscopy. *Langmuir* **2012**, *28*, 16738–16744. doi:10.1021/la303891j.
19. Pfreundschuh, M.; Alsteens, D.; Hilbert, M.; Steinmetz, M.O.; Müller, D.J. Localizing chemical groups while imaging single native proteins by high-resolution atomic force microscopy. *Nano Letters* **2014**, *14*, 2957–2964. doi:10.1021/nl5012905.
20. Heu, C.; Berquand, A.; Elie-Caille, C.; Nicod, L. Glyphosate-induced stiffening of HaCaT keratinocytes, a Peak Force Tapping study on living cells. *Journal of Structural Biology* **2012**, *178*, 1–7. doi:10.1016/J.JSB.2012.02.007.
21. Calzado-Martín, A.; Encinar, M.; Tamayo, J.; Calleja, M.; San Paulo, A. Effect of Actin Organization on the Stiffness of Living Breast Cancer Cells Revealed by Peak-Force Modulation Atomic Force Microscopy. *ACS Nano* **2016**, *10*, 3365–3374. doi:10.1021/acs.nano.5b07162.

22. Alsteens, D.; Trabelsi, H.; Soumillion, P.; Dufrène, Y.F. Multiparametric atomic force microscopy imaging of single bacteriophages extruding from living bacteria. *Nature Communications* **2013**, *4*, 2926. doi:10.1038/ncomms3926.
23. Schillers, H.; Medalsy, I.; Hu, S.; Slade, A.L.; Shaw, J.E. PeakForce Tapping resolves individual microvilli on living cells. *Journal of Molecular Recognition* **2016**, *29*, 95–101. doi:10.1002/jmr.2510also.
24. Nievergelt, A.P.; Erickson, B.W.; Hosseini, N.; Adams, J.D.; Fantner, G.E. Studying biological membranes with extended range high-speed atomic force microscopy. *Scientific Reports* **2015**, *5*, 11987. doi:10.1038/srep11987.
25. Eskandarian, H.H.A.; Odermatt, P.D.P.; Ven, J.J.X.Y.; Hannebelle, M.T.M.; Nievergelt, A.A.P.; Dhar, N.; McKinney, J.J.D.; Fantner, G.E.G. Division site selection linked to inherited cell surface wave troughs in mycobacteria. *Nature Microbiology* **2017**, *2*, 1–6. doi:10.1038/nmicrobiol.2017.94.
26. Ratcliff, G.C.; Erie, D.a.; Superfine, R. Photothermal modulation for oscillating mode atomic force microscopy in solution. *Applied Physics Letters* **1998**, *72*, 1911–1913. doi:10.1063/1.121224.
27. Ramos, D.; Tamayo, J.; Mertens, J.; Calleja, M. Photothermal excitation of microcantilevers in liquids. *Journal of Applied Physics* **2006**, *99*, 124904. doi:10.1063/1.2205409.
28. Labuda, A.; Kobayashi, K.; Kiracofe, D.; Suzuki, K.; Grütter, P.H.; Yamada, H. Comparison of photothermal and piezoacoustic excitation methods for frequency and phase modulation atomic force microscopy in liquid environments. *AIP Advances* **2011**, *1*, 022136. doi:10.1063/1.3601872.
29. Nievergelt, A.P.; Adams, J.D.; Odermatt, P.D.; Fantner, G.E. High-frequency multimodal atomic force microscopy. *Beilstein Journal of Nanotechnology* **2014**, *5*, 2459–2467. doi:10.3762/bjnano.5.255.
30. Nievergelt, A.P.; Andany, S.H.; Adams, J.D.; Hannebelle, M.T.; Fantner, G.E. Components for high-speed atomic force microscopy optimized for low phase-lag. 2017 IEEE International Conference on Advanced Intelligent Mechatronics (AIM). IEEE, 2017, pp. 731–736. doi:10.1109/AIM.2017.8014104.
31. Nievergelt, A.P.; Banterle, N.; Andany, S.H.; Gönczy, P.; Fantner, G.E. High-speed photothermal off-resonance atomic force microscopy reveals assembly routes of centriolar scaffold protein SAS-6. *Nature Nanotechnology* **2018**, *13*. doi:10.1038/s41565-018-0149-4.
32. Machlus, K.R.; Thon, J.N.; Italiano, J.E. Interpreting the developmental dance of the megakaryocyte: a review of the cellular and molecular processes mediating platelet formation. *British Journal of Haematology* **2014**, *165*, 227–236. doi:10.1111/bjh.12758.
33. Jain, N.C. A scanning electron microscopic study of platelets of certain animal species. *Thrombosis et diathesis haemorrhagica* **1975**, *33*, 501–7.
34. Jones, M.I.; McColl, I.R.; Grant, D.M.; Parker, K.G.; Parker, T.L. Protein adsorption and platelet attachment and activation, on TiN, TiC, and DLC coatings on titanium for cardiovascular applications. *Journal of Biomedical Materials Research* **2000**, *52*, 413–421. doi:10.1002/1097-4636(200011)52:2<413::AID-JBM23>3.0.CO;2-U.
35. Fantner, G.E.; Barbero, R.J.; Gray, D.S.; Belcher, A.M. Kinetics of antimicrobial peptide activity measured on individual bacterial cells using high-speed atomic force microscopy. *Nature nanotechnology* **2010**, *5*, 280–5. doi:10.1038/nnano.2010.29.
36. Peric, O.; Hannebelle, M.; Adams, J.D.; Fantner, G.E. Microfluidic bacterial traps for simultaneous fluorescence and atomic force microscopy. *Nano Research* **2017**, *10*, 3896–3908. doi:10.1007/s12274-017-1604-5.
37. And, R.D.B.; Verran\*, J.; And, M.V.J.; Bhakoo, M. Use of the Atomic Force Microscope To Determine the Effect of Substratum Surface Topography on Bacterial Adhesion **2002**. doi:10.1021/LA011142P.
38. Pistolesi, S.; Pogni, R.; Feix, J.B. Membrane Insertion and Bilayer Perturbation by Antimicrobial Peptide CM15. *Biophysical Journal* **2007**, *93*, 1651–1660. doi:10.1529/biophysj.107.104034.
39. Graham, L.L.; Beveridge, T.J. Structural differentiation of the *Bacillus subtilis* 168 cell wall. *Journal of bacteriology* **1994**, *176*, 1413–21.
40. Mercier, R.; Kawai, Y.; Errington, J. Excess Membrane Synthesis Drives a Primitive Mode of Cell Proliferation. *Cell* **2013**, *152*, 997–1007. doi:10.1016/j.cell.2013.01.043.
41. Utke, I.; Hoffmann, P.; Melngailis, J. Gas-assisted focused electron beam and ion beam processing and fabrication. *Journal of Vacuum Science & Technology B: Microelectronics and Nanometer Structures* **2008**, *26*, 1197. doi:10.1116/1.2955728.

42. Huth, M.; Porрати, F.; Schwalb, C.; Winhold, M.; Sachser, R.; Dukic, M.; Adams, J.; Fantner, G. Focused electron beam induced deposition: A perspective. *Beilstein Journal of Nanotechnology* **2012**, *3*, 597–619. doi:10.3762/bjnano.3.70.
43. Uchihashi, T.; Kodera, N.; Ando, T. Guide to video recording of structure dynamics and dynamic processes of proteins by high-speed atomic force microscopy. *Nature Protocols* **2012**, *7*, 1193–1206. doi:10.1038/nprot.2012.047.
44. Adams, J.D.; Nievergelt, A.; Erickson, B.W.; Yang, C.; Dukic, M.; Fantner, G.E. High-speed imaging upgrade for a standard sample scanning atomic force microscope using small cantilevers. *Review of Scientific Instruments* **2014**, *85*, 093702. doi:10.1063/1.4895460.