

Title: Clinical Experience with the Fifth-generation of a Breast Implant with a Smooth, Fine Surface from a Korean Manufacturer in Asian Women

Running head: Clinical experience with the BellaGel[®] SmoothFine

Dong Yeon Hwang, M.D.,^{1*} Sang Wha Kim, M.D., Ph.D.,² Joon Yong Jung, M.D.,³

¹NANA Plastic Surgery, Seoul, Korea, Seoul, Korea,

²Department of Plastic and Reconstructive Surgery,

Seoul National University College of Medicine,

Seoul National University Hospital, Seoul, Korea,

³Department of Radiology, The Catholic University of Korea College of Medicine,

Seoul St. Mary's Hospital, Seoul, Korea,

***Address correspondence to:** Dong Yeon Hwang, M.D.,

Current affiliation: NANA Plastic Surgery,

Address: 492 Gangnam-daero, Gangnam-gu, Seoul 06119 Korea,

Tel: +82-2-544-0601,

Fax: +82-2-549-0601,

E-mail: wannabe0@naver.com.

ABSTRACT

Background In this study, we describe our clinical experience with the fifth-generation of a breast implant with a smooth, fine surface from a Korean manufacturer (BellaGel[®] SmoothFine; HansBiomed Co. Ltd., Seoul, Korea) in Asian women.

Methods We analyzed 223 women (mean age=35.28±9.45 years and mean follow-up period=12.03±2.48 months), comprising 118 bilateral cases and 109 unilateral ones, who received breast augmentation using the BellaGel[®] SmoothFine at our hospital between June 4, 2018 and February 28, 2019. For safety assessment, we analyzed frequencies of postoperative complications and overall survival of the BellaGel[®] SmoothFine.

Results Postoperatively, complications (12 cases, 5.38%) include asymmetry (3 cases, 1.35%), hematoma (2 cases, 0.90%), hypertrophic scars (2 cases, 0.90%), wound disruption (2 cases, 0.90%), rippling (1 case, 0.45%), capsular contracture (1 case, 0.45%), stretch deformities with skin excess (1 case, 0.45%). In addition, time-to-events were calculated as 10.94±0.64 months (95% CI 9.69-12.19).

Conclusions Here, we describe our clinical experience with the BellaGel[®] SmoothFine. Our results are of significance in that this is the first report about the fifth-generation of a breast implant with a smooth, fine surface from a Korean manufacturer in Asian women.

Keywords: Clinical Study; Retrospective Studies; Surgical Procedures, Operative; Postoperative Period; Follow-Up Studies.

Introduction

In the US in 2018, breast augmentation using 313,735 implants was performed. Of these, 29,236 and 19,149 were removed from patients following aesthetic and reconstructive surgeries, respectively [1]. Nevertheless, such patients are vulnerable to postoperative complications, including breast implant-associated anaplastic large cell lymphoma (BIA-ALCL). It is an extremely rare, fatal condition with a causal relationship with a textured device. In addition, it is characterized by formation of scar tissue, known as a tissue capsule, around a device; it occurs at approximately 7-10 years postoperatively on average [2-4]. In 2018, de Boer et al. reported that there was an age-dependent increase in its risk in women receiving a breast implant; it was estimated at 1/35,000, 1/12,000 and 1/7,000 in patients aged 50, 70 and 75, respectively [5]. Nevertheless, a lack of safety studies in the US remains problematic; two representative breast implant manufacturers received a warning letter from the Food and Drug Administration (FDA) in March of 2019 because they failed to conduct long-term follow-up studies required for the regulatory approval [6]. Therefore, safety studies evaluating breast augmentation using an implant in Korea are also warranted [7].

The BellaGel[®] SmoothFine (HansBiomed Co. Ltd., Seoul, Korea) is the fifth-generation of a breast implant with a smooth, fine surface from a Korean manufacturer. Its manufacturing process entails a rigorous analysis of physical features of Korean women [8].

The BellaGel[®] implants, including the BellaGel[®] SmoothFine, have been exported to 30 countries worldwide since they were first developed in 2005 [9]. It is noteworthy that their efficacy and safety have been established through evidence-based studies [8-12].

As recently described, it was found to be more excellent than its competitors in Korea [11].

Here, we describe our clinical experience with the BellaGel[®] SmoothFine in Asian women.

Patients and methods

Study design

We evaluated a total of 223 women (446 breasts) who received breast augmentation using the BellaGel[®] SmoothFine and were followed up for more than one year at our hospital between June 4, 2018 and February 28, 2019. We adhered to the relevant ethics guidelines and the Declaration of Helsinki. The current study obtained the approval from the Internal Institutional Review Board of the Korea National Institute of Bioethics Policy in November 2, 2020 (approval # P01-202010-19-016). Written informed consent was waived due to its retrospective nature.

Study material

The BellaGel[®] SmoothFine is manufactured through a process where the mandrel surface is treated with sandblast and its fine structure is transferred to the surface of shell. The mobility of cell and tissue varies depending on the surface topography. This is associated with variability in the occurrence of complications of breast augmentation using an implant [13]. It is equipped with a softness *as well as* a refined, smooth surface with a roughness of 5.96 μm , which is a different feature from traditional smooth

surface, according to the International Organization for Standardization (ISO) 14607 Annex H Test for surface characteristics (Fig. 1 and 2) [14]. It is covered with five layers of shell, within which there is a barrier layer that efficiently prevents the leakage of a gel due to a rupture (Fig. 3). In addition, it is equipped with a round shape, a high degree of viscoelasticity and excellent gel properties; it is advantageous in creating a natural breast silhouette (Fig. 4). A previous study reported that the BellaGel® SmoothFine showed an almost 10-fold lower incidence of complications than its competitors [10].

Treatment protocol

Evidence-based treatment protocol for breast augmentation using the BellaGel® SmoothFine has been previously described in detail [8,11]. Our treatment protocol is schematically illustrated in Fig. 5.

Postoperatively, the patients were regularly followed up at a 1-week interval for the first three weeks, a 3-month interval for the next one year and thereafter. In addition, they were also examined on magnetic resonance imaging (MRI) studies at 3 years after surgery and at a 2-year interval thereafter, as recommended by the FDA [8,11].

Criteria for assessing the patients' postoperative course and analyzing their data

The patients' demographic and clinical characteristics were analyzed using their medical records, as previously described [8,11]. In more detail, we assessed the safety based on frequencies of postoperative complications and survival of the BellaGel® SmoothFine, as previously described [8,11].

After presenting the data as mean±standard deviation or standard error or the number

with percentage, we performed a data analysis with the SPSS ver. 23 (IBM Corp., Armonk, NY). Statistical significance was defined as $P < 0.05$. The Kaplan-Meier survival was estimated.

Results

Demographic and clinical characteristics of the patients

In the current study, 223 women (446 breasts) were evaluated; they comprise 118 bilateral cases and 109 unilateral ones. As presented in Table 1, their mean age was 35.28 ± 9.45 years (range, 23-47) and mean follow-up period was 12.03 ± 2.48 months.

Illustrative cases are shown in Fig. 6 and 7.

Safety outcomes

Postoperatively, complications (12 cases, 5.38%) include asymmetry (3 cases, 1.35%), hematoma (2 cases, 0.90%), hypertrophic scars (2 cases, 0.90%), wound disruption (2 cases, 0.90%), rippling (1 case, 0.45%), capsular contracture (1 case, 0.45%), stretch deformities with skin excess (1 case, 0.45%) (Table 2).

In addition, time-to-events (TTEs) were calculated as 10.94 ± 0.64 months (95% CI 9.69-12.19) (Table 3). Furthermore, survival and hazards are shown in Table 4 and Fig. 8 and 9.

Discussion

To date, anatomical implants have become available to maximally imitate a natural shape of the breast by providing a more fullness in the lower pole. But their availability has been greatly decreased. Currently, a balance between the width and volume of a breast implant therefore serves as a determinant of aesthetic outcomes of a silicone gel-filled breast implant [15,16].

Korean women with a breast width of approximately 11-12 cm require both a full B- or C-cup size and a natural feel [10,17]. The BellaGel® SmoothFine is a silicone gel-filled breast implant whose manufacturing process encompasses the consideration of physical characteristics of Korean women, such as the size of chest and skin features, and their westernized life style and then compensates the disadvantages of other brands of the breast implant [8].

The BellaGel® SmoothFine is also the best choice for Korean women who are characterized by a smaller amount of the breast parenchyma and more elastic skin than Caucasian women [8]. This is based on previous published studies showing that both the size of breast and skin tonicity are two determinants of aesthetic outcomes of breast augmentation using an implant [18,19].

Conventional types of anatomical breast implants equipped with a thick shell and a highly-viscous silicone gel are vulnerable to decreased tensile strength of the shell despite its thickness; they do not adjust to the stress due to continuous exercise. This may lead to fracture due to the increased fatigue [20,21]. It can therefore be inferred that women receiving anatomical implants are at increased risks of derangement in the integrity of a silicone gel and the resulting shape deformity [22,23]. In addition, a breast implant with a thick shell and a soft silicone gel reveals its disadvantages such as an

insufficient level of tensile strength [24,25]. Therefore, such breast implants cannot retain their integrity in a standing posture after a considerable length of time postoperatively. This may cause serious problems in women with ptotic or elastic breast with thin overlying skin. Tensile force and strength of the shell as well as the softness and viscosity of a silicone gel are therefore key determinants of biomechanical properties of a breast implant [25].

To summarize, we found a total of 12 cases (5.38%) of postoperative complications. We also showed that TTEs were calculated as 10.94 ± 0.64 months (95% CI 9.69-12.19).

Our results cannot be generally interpreted due to the following reasons:

- (1) We failed to clarify a relationship between physico-chemical properties and surface characteristics of the BellaGel® SmoothFine and its short-term safety.
- (2) We failed to identify associations of rheological and biomechanical properties of the BellaGel® SmoothFine with westernized life style of Korean women who are much interested in pilates, Yoga or fitness.
- (3) We could not completely rule out the possibility of a study design bias because this is a manufacturer-sponsored, retrospective study [11].

Conclusions

In conclusion, we describe our clinical experience with breast augmentation using the BellaGel® SmoothFine. Our results are of significance in that this is the first report about the fifth-generation of a breast implant with a smooth, fine surface from a Korean manufacturer in Asian women.

Funding

The current study was sponsored by the HansBiomed Co. Ltd.

Conflict of interest

The authors have nothing to declare in relation to the current work.

References

1. American Society of Plastic Surgeons. The 2018 Plastic Surgery Report.
<https://www.plasticsurgery.org/documents/News/Statistics/2018/plastic-surgery-statistics-report-2018.pdf>. Accessed September 25, 2019.
2. Hamdi M (2019) Association Between Breast Implant-Associated Anaplastic Large Cell Lymphoma (BIA-ALCL) Risk and Polyurethane Breast Implants: Clinical Evidence and European Perspective. *Aesthet Surg J* 39: S49-S54.
3. Cardoso MJ, Wyld L, Rubio IT, Leidenius M, Curigliano G, et al. (2019) EUSOMA position regarding breast implant associated anaplastic large cell lymphoma (BIA-ALCL) and the use of textured implants. *Breast* 44: 90-93.
4. Collett DJ, Rakhorst H, Lennox P, Magnusson M, Cooter R, et al. (2019) Current Risk Estimate of Breast Implant-Associated Anaplastic Large Cell Lymphoma in Textured Breast Implants. *Plast Reconstr Surg* 143: 30S-40S.
5. de Boer M, van Leeuwen FE, Hauptmann M, Overbeek LIH, de Boer JP, et al. (2018) Breast Implants and the Risk of Anaplastic Large-Cell Lymphoma in the Breast. *JAMA Oncol* 4: 335-341.
6. US Food and Drug Administration. FDA issues warning letters to two breast implant manufacturers for failure to comply with post-approval study

requirements. <https://www.fda.gov/news-events/press-announcements/fda-issues-warning-letters-two-breast-implant-manufacturers-failure-comply-post-approval-study>. Accessed September 25, 2019.

7. Kim JH, Paik NS, Nam SY, Cho Y, Park HK (2020) The Emerging Crisis of Stakeholders in Implant-based Augmentation Mammoplasty in Korea. *J Korean Med Sci* 35: e103.
8. Sung JY, Jeong JP, Moon DS, Kim MS, Kim HC, et al. (2019) Short-term Safety of Augmentation Mammoplasty Using the BellaGel Implants in Korean Women. *Plast Reconstr Surg Glob Open* 7: e2566.
9. Choi MS, Chang JH, Seul CH (2020) A multi-center, retrospective, preliminary observational study to assess the safety of BellaGel® after augmentation mammoplasty. *Eur J Plast Surg* 43: 577-582.
10. Han J, Jeong JH, Bang SI, Heo CY (2019) BellaGel breast implant: 4-year results of a prospective cohort study. *J Plast Surg Hand Surg* 53: 232-239.
11. Yoon S, Chang JH (2020) Short-term Safety of a Silicone Gel-filled Breast Implant: A Manufacturer-sponsored, Retrospective Study. *Plast Reconstr Surg Glob Open* 8: e2807.
12. Oh JS, Jeong JH, Myung Y, Oh J, Kang SH, et al. (2020) BellaGel breast

- implant: 6-Year results of a prospective cohort study. *Arch Plast Surg* 47: 235-241.
13. Valencia-Lazcano AA, Alonso-Rasgado T, Bayat A (2013) Characterisation of breast implant surfaces and correlation with fibroblast adhesion. *J Mech Behav Biomed Mater* 21: 133-148.
14. Frame (2017) The waterfall effect in breast augmentation. *Gland Surg* 6: 193-202.
15. Gardani M, Bertozzi N, Grieco MP, Pesce M, Simonacci F, et al. (2017) Breast reconstruction with anatomical implants: A review of indications and techniques based on current literature. *Ann Med Surg* 21: 96-104.
16. Yi KH, Sohn BH (2017) Comparison of Breast Measurement Ratios Before and After Breast Augmentation Using Photogrammetric Ratio Measurements (PRM). *J Ergon Soc Korea* 36: 305-323.
17. Khan UD (2007) Muscle-splitting breast augmentation: a new pocket in a different plane. *Aesthetic Plast Surg* 31: 553-558.
18. Huang GJ, Wichmann JL, Mills DC (2011) Transaxillary subpectoral augmentation mammoplasty: a single surgeon's 20-year experience. *Aesthet Surg J* 31: 781-801.

19. Maxwell GP, Gabriel A (2017) Breast implant design. *Gland Surg* 6: 148-153.
20. Takayanagi S (2012) Augmentation mammoplasty using implants: a review. *Arch Plast Surg* 39: 448-451.
21. Govrin-Yehudain J, Dvir H, Preise D, Govrin-Yehudain O, Govreen-Segal D (2015) Lightweight breast implants: a novel solution for breast augmentation and reconstruction mammoplasty. *Aesthet Surg J* 35: 965-971.
22. O'Shaughnessy K (2015) Evolution and update on current devices for prosthetic breast reconstruction. *Gland Surg* 4: 97-110.
23. Jewell ML, Bengtson BP, Smither K, Nuti G, Perry T (2019) Physical Properties of Silicone Gel Breast Implants. *Aesthet Surg J* 39: 264-275.
24. Teck Lim G, Valente SA, Hart-Spicer CR, Evancho-Chapman MM, Puskas JE, et al. (2013) New biomaterial as a promising alternative to silicone breast implants. *J Mech Behav Biomed Mater* 21: 47-56.
25. Spear SL, Davison SP (2003) Aesthetic subunits of the breast. *Plast Reconstr Surg* 112: 440-447.

Tables

Table 1. Baseline characteristics of the patients (n=223; 446 breasts).

Variables		Values
Age (years)		35.28±9.45 (23-47)
Follow-up period (months)		12.03±2.48 (12-16)
Sex		
	Men	0 (0.00%)
	Women	223 (100.00%)
Height (cm)		162.92±20.78 (158-173)
Weight (kg)		51.37±5.84 (46-68)
Operation time (hours)		1.09±0.38 (0.75-2.00)
Round of surgery		
	Primary augmentation mammoplasty	214 (95.96%)
	Revision augmentation mammoplasty	9 (4.04%)
Profile and volume of breast implant (cc)		
	Right side (n=223)	
	H275	2 (0.90%)
	H300	4 (1.79%)
	H325	16 (7.17%)
	H350	10 (4.48%)
	H375	3 (1.35%)
	H400	3 (1.35%)

	M150	1 (0.45%)
	M225	4 (1.79%)
	M250	12 (5.38%)
	M275	43 (19.28%)
	M300	48 (21.52%)
	M325	58 (26.01%)
	M350	13 (5.83%)
Left side (n=105)*		
	H275	2 (0.90%)
	H300	4 (1.79%)
	H325	10 (4.48%)
	H350	11 (4.93%)
	H375	3 (1.35%)
	M225	2 (0.90%)
	M250	9 (4.04%)
	M275	14 (6.28%)
	M300	26 (11.66%)
	M325	18 (8.07%)
	M350	5 (2.24%)
	M375	1 (0.45%)
Type of surgical methods		
	Dual-plane augmentation mastopexy	6 (2.69%)
	Dual-plane technique	211 (94.62%)

	Neo-subpectoral technique	3 (1.35%)
	Neo-subpectotal technique with ADM	1 (0.45%)
	Subglandular technique	2 (0.89%)
Type of surgical incision		
	Axillary incision	148 (66.37%)
	IMF incision	75 (33.63%)

Abbreviations: ADM, acellular dermal matrix; IMF, inframammary fold.

Values are mean±standard deviation with the range in parenthesis or the number of the patients with percentage, where appropriate

*The remaining 118 patients received other brands of a breast implant on the left side for correction of the asymmetry.

Table 2. Postoperative complications.

Variables	Values
Asymmetry	3 (1.35%)
Hematoma	2 (0.90%)
Hypertrophic scars	2 (0.90%)
Wound disruption	2 (0.90%)
Rippling	1 (0.45%)
Capsular contracture	1 (0.45%)
Stretch deformities with skin excess	1 (0.45%)

Values are the number of the patients with percentage.

Table 3. Overall complication-free survival.

N	n	Censored value	Time-to-events (months)	95% CI
223	12	211 (94.62%)	10.94±0.64	9.69-12.19

Note: N, total number of cases; n, incidences of postoperative complications; CI,

confidence intervals.

Values are mean±standard error or the number of the patients with percentage, where appropriate.

Table 4. Cumulative complication-free survival.

FU (months)	N	n	Survival rate	95% CI
1	223	6	0.973±0.011	0.952-0.995
5	16	1	0.912±0.060	0.802-1.000
6	14	1	0.847±0.084	0.698-1.000
10	7	1	0.726±0.133	0.507-1.000
12	5	3	0.290±0.168	0.094-0.901

Note: FU, follow-up; N, total number of cases; n, incidences of postoperative complications; CI, confidence intervals.

Values are mean±standard error.

Figure legends

Fig. 1. The BellaGel® SmoothFine based on the International Organization for Standardization classification.

The BellaGel® SmoothFine is equipped with a refined, smooth surface with a roughness of 5.96 μm , which is a different feature from traditional smooth surface.

Fig. 2. Surface characteristics of the BellaGel® SmoothFine.

The surface of the BellaGel® SmoothFine was coated with platinum for 60 seconds and then examined using a scanning electron microscope (MIRA3 LM; TESCAN, Kohoutovice, Czech) at magnifications of (A) 50 \times and (B) 200 \times .

Fig. 3. The structure of the BellaGel® SmoothFine.

The BellaGel SmoothFine® is covered with a five layers of shell, within which there is a barrier layer that efficiently prevents the leakage of a gel due to a rupture.

Fig. 4. The advantages of the BellaGel® SmoothFine.

The BellaGel® SmoothFine has a round surface, a high degree of viscoelasticity and excellent gel properties; it is advantageous in creating a natural breast silhouette.

Fig. 5. A multi-disciplinary approach to an implant-based augmentation mammoplasty using the BellaGel® SmoothFine.

A multi-disciplinary approach to an implant-based augmentation mammoplasty using the BellaGel® SmoothFine was developed for the purposes of maximizing its safety outcomes.

Fig. 6. Primary augmentation mammoplasty using the BellaGel® SmoothFine.

(A) Preoperatively, a 25-year-old woman visited us with a chief complaint of small breast size. The patient received the BellaGel® SmoothFine at a volume of 350 and 325 cc for the right and left sides, respectively, *via* an inframammary fold incision using a modified dual-plane type 1 technique. (B) At 1 month postoperatively, the patient achieved an improvement in the size of the breast. (C) At 3 months postoperatively, the patient was satisfied with the external appearance of the breast.

Fig. 7. Reoperation using the BellaGel® SmoothFine.

(A) Preoperatively, a 34-year-old woman had a 7-year-history of receiving an 300-cc implant of unknown brand *via* an axillary incision. Due to dissatisfaction with the size and shape of the breast, the patient underwent inferolateral capsulorrhaphy with superomedial “mirror image” capsulotomy. The pre-existing implant was replaced with the BellaGel® SmoothFine at a volume of 350 cc for both sides. (B) At 1 month postoperatively, the patient was satisfied with the external appearance of the breast.

Fig. 8. Kaplan-Meier cumulative survival.

In our series, time-to-events were calculated as 10.94 ± 0.64 months (95% CI 9.69-12.19).

Fig. 9. Kaplan-Meier cumulative hazards.